

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
1,1	DCU	LIA P	AF4.1.1 : To collect the analogue data coming from FPU, to demodulate them and to filter them	Loss of PSD + filter (among 32) of a LIA P board (electronic or connector failure)	Loss of a photometric scientific data sent to the DAQ (degraded IF1.1)	Reception by the DAQ of one of the 32 analogue channels degraded for treatment (digitalisation) and sending to the DPU (degraded IF1.1)	Loss of a photometric scientific data among a potential of 9*32 data (degraded IF1)	9 photometric LIA board Detection of the loss of the data during the analysis by the scientists	3	
1,1 Bis	DCU	LIA S	AF4.1.1 : To collect the analogue data coming from FPU, to demodulate them and to filter them	Loss of PSD + filter (among 24) of a LIA board (electronic or connector failure)	Loss of a spectrometric scientific data sent to the DAQ (degraded IF1.1)	Reception by the DAQ of one of the 24 analogue channels degraded for treatment (digitalisation) and sending to the DPU (degraded IF1.1)	Loss of a spectrometric scientific data among a potential of 3*24 data (degraded IF1)	3 spectrometric LIA board Detection of the loss of the data during the analysis by the scientists	3	
1,2		LIA P	AF4.1.2 : To multiplex and transfer the analogue data	Failure of a multiplexer 16 > 1 (among 2) of a LIA P board	Loss of 16 photometric scientific data sent to the DAQ (degraded IF1.1)	Loss of one of the 32 analogue channels (not sent to the DPU) (degraded IF1.1)	Loss of 16 photometric scientific data among a potential of 9*32 (degraded IF1)	2 multiplexers 16 > 1 per LIA P board Detection of the loss of the 16 data during the analysis by the scientists If the 3 chains of T/C measure (temperature) are used for the regulation of the bolometers (via control of the cooler), the distribution of the T/C must be such as the loss of a multiplexer 16 > 1 does not lead to the loss of the 3 chains of temperature. This distribution implies 3 distinct analogue channels towards the multiplexers 4 > 1 of DAQ module.	3	
1,2 Bis		LIA S	AF4.1.2 : To multiplex and transfer the analogue data	Failure of a multiplexer 12 > 1 (among 2) of a LIA board	Loss of 12 spectrometric scientific data sent to the DAQ (degraded IF1.1)	Loss of one of the 24 analogue channels (not sent to the DPU) (degraded IF1.1)	Loss of 12 spectrometric scientific data among a potential of 3*24 data (degraded IF1)	2 multiplexers 12 > 1 per LIA S board Detection of the loss of the 12 data during the analysis by the scientists	3	
1,3	DCU	LIA P	AF4.1.1 : To collect the analogue data coming from FPU, to demodulate them and to filter them AF4.1.2 : To multiplex and transfer the analogue data	Short circuit of a component Heating of a LIA P board following internal failure	Loss of 32 photometric scientific data sent to the DAQ (degraded IF1.1)	Loss of 2 of the 32 analogue channels (not sent to the DPU) (degraded IF1.1)	Loss of 32 photometric scientific data among a potential of 9*32 data (degraded IF1) Risk propagation of failure (PoF) towards the other LIA board by thermal propagation (degradation of the other LIA board) total Loss of the mission Photometer (loss of all the scientific data)	The LIA board are not redundant and cannot be controlled ON/OFF individually. The heating of a LIA board is detected by a specific temperature measurement (Housekeeping parameters) returned towards the DPU. If the case of a risk of failure propagation towards the other LIA boards leading to a total loss of the scientific data is confirmed, each analogue module (LIA) must be individually commandable ON/OFF by means of low level commands. If the 3 chains of T/C measure (temperature) are used for the regulation of the bolometers (via control of the cooler), the distribution of the T/C must be such as the loss of a LIA board does not lead to the loss of the 3 chains of temperature.	3	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
1,3 Bis	DCU	LIA S	AF4.1.1 : Collect the analogue data coming from FPU, demodulate them and filter them AF4.1.2 : Multiplex and transfer the analogue data	Short circuit component Heating of a LIA board following internal failure	Loss of 24 spectrometric scientific data sent to the DAQ (degraded IF1.1)	Loss of 2 of the 24 analogue channels (not sent to the DPU) (degraded IF1.1)	Loss of 24 spectrometric scientific data among a potential of 3*24 data (degraded IF1) Risk of propagation of failure (PoF) towards the other LIA board by thermal propagation (degradation of the other LIA board). Total loss of the spectrometric mission (loss of all the scientific data)	The LIA board are not redundant and cannot be controlled ON/OFF individually. The heating of a LIA board is detected by a specific temperature measurement (Housekeeping parameters) returned towards the DPU. If the case of a risk of failure propagation towards the other LIA boards leading to a total loss of the scientific data is confirmed, each analogue module (LIA) must be individually commandable ON/OFF by means of low level commands.	3	
2,1	DCU	BIAS	AF 4.2.2 : To adapt and transfer the analogue controls data (signal of modulation) towards the photometric or spectrometric detectors	Failure of one of the 4 Bias DAC dedicated to the photometric detectors: no transfer no adaptation of the signal (amplitude, frequency, gain)	No generation of corresponding analogue controls (degraded IF1.2.2)	No sending of analogue control to the photometric detectors of the FPU (degraded IF1.2)	Loss of all the data from the photometric detectors of the FPU towards the DPU (lost IF1) Need for switching on the redundant chain of the DRCU (except LIA) to recover all detectors data	Detection of the loss of the data during the analysis by the scientists The switching is carried out by the switching of supply " Primary power line " (28 V) at the satellite level. The switching procedure must be performed safely (shut off the blocks before switching of power supply). Potential problems with the Bistable relays.	3R	
2,1 Bis	DCU	BIAS	AF 4.2.2 : To adapt and transfer the analogue controls data (signal of modulation) towards the photometric or spectrometric detectors	Failure of one of the 2 Bias DAC dedicated to the spectrometric detectors: no transfer no adaptation of the signal (amplitude, frequency, gain)	No generation of corresponding analogue controls (degraded IF1.2.2)	No sending of analogue control to the spectrometric detectors of the FPU (degraded IF1.2)	Loss of all the data from the spectrometric detector of the FPU towards the DPU (lost IF1) Need for switching on the redundant chain of the DRCU (except LIA) to recover all detectors data	Detection of the loss of the data during the analysis by the scientists The switching is carried out by the switching of supply " Primary power line " (28 V) at the satellite level. The switching procedure must be performed safely (shut off the blocks before switching of power supply). Potential problems with the Bistable relays.	3R	
2,2		BIAS	AF 4.2.2 : To adapt and transfer the analogue controls data (signal of modulation) towards the photometric or spectrometric detectors	Failure of one of the 4 Bias DAC dedicated to the photometric detectors: bad adaptation (amplitude), bad interpretation of the amplitude parameter emitted by the DAQ - I/F	No adaptation of the sinusoidal signal emitted by Sine DAC (amplitude of the signal corresponding to that of Sine DAC)	Transmission of a signal not adapted to the photometric detectors of the FPU (degraded IF1.2)	Loss of all the data from the photometric detectors of the FPU towards the DPU Risk of deterioration of the corresponding detectors (worst case; non compatible amplitude; double failure) (lost IF1)	Amplitude parameters sent by the DAQ - I/F in each Bias Dac are returned to the DPU (Housekeeping parameters). The value of amplitude issued from Bias DAC is not a Housekeeping parameter. In the case of the non adaptation of the signal is not compatible with detectors and lead to their deterioration, the second reading of the value of amplitude issued from Bias Dac and returned to DAQ I/F must lead to re-sent the nominal amplitude parameter towards Bias DAC at fault.	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
2.2 Bis		BIAS	AF 4.2.2 : To adapt and transfer the analogue controls data (signal of modulation) towards the photometric or spectrometric detectors	Failure of one of the 2 Bias DAC dedicated to the spectrometric detectors: bad adaptation (amplitude), bad interpretation of the amplitude parameter emitted by the DAQ - I/F	No adaptation of the sinusoidal signal emitted by Sine DAC (amplitude of the signal corresponding to that of Sine DAC)	Transmission of a signal not adapted on the spectrometric detectors of the FPU (degraded IF1.2)	Loss of all the data from the spectrometric detectors of the FPU towards the DPU Risk of deterioration of the corresponding detectors (worst case; non compatible amplitude: double failure) (lost IF1)	Amplitude parameters sent by the DAQ - I/F in each Bias Dac are returned to the DPU (Housekeeping parameters). The value of amplitude issued from Bias DAC is not a Housekeeping parameter. In the case of the non adaptation of the signal is not compatible with detectors and lead to their deterioration, the second reading of the value of amplitude issued from Bias Dac and returned to DAQ I/F must lead to re-sent the nominal amplitude parameter towards Bias DAC at fault.	3R	
2.3		BIAS	DF 3 : To ensure the sorting of information of control coming from DAQ	Failure of Bias Logic (no sorting, no transmission) dedicated to the photometric detectors	No generation of numerical controls relating to photometric detections towards: Sine Dac for modulation signal towards the detectors Bias DAC for the amplitude parameters Heater DACs for heating of the JFET box when starting JFET DACs for power supply of the JFET (lost IF1.2.2)	No sending towards the FPU of the analogue controls of modulation signal towards the detectors, value of amplitude, heating of the JFET box and power supply of the JFET (degraded IF1.2)	Loss of all the data from the photometric detectors of the FPU towards the DPU (lost IF1) Need for switching on the redundant chain of the DRCU (except LIA) to recover all detectors data	Idem item 2.1	3R	
2.3 Bis		BIAS	DF 3 : To ensure the sorting of information of control coming from DAQ	Failure of Bias Logic (no sorting, no transmission) dedicated to the spectrometric detectors	No generation of numerical controls relating to photometric detections towards: Sine Dac for modulation signal towards the detectors Bias DAC for the amplitude parameters Heater DACs for heating of the JFET box when starting JFET DACs for power supply of the JFET (lost IF1.2.2)	No sending towards the FPU of the analogue controls of signal of modulation towards the detectors, value of amplitude, heating of the JFET box and power supply of the JFET (degraded IF1.2)	Loss of all the data from the spectrometric detectors of the FPU towards the DPU (lost IF1) Need for switching on the redundant chain of the DRCU (except LIA) to recover all detectors data	Idem item 2.1	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
2.4		BIAS	DF 2 : To transfer the controlling of specific amplitude to each Bias DAC	Failure of an output of Bias Logic: no transmission of an amplitude parameter relating to the signal of modulation of the photometric detectors	No generation of a numerical control relating to the photometric detectors towards Bias DAC (IF 1.2.2 degraded)	No sending towards the FPU of an analogue control of modulation signal towards the photometric detectors (degraded IF1.2)	Loss of all the data from the photometric detectors of the FPU towards the DPU (lost IF1) Need for switching on the redundant chain of the DRCU (except LIA) to recover all detectors data	DAQ I/F transfers to each Bias DAC via Bias Logic, the value of amplitude of the signal to be generated for the modulation of the photometric detectors. It is considered that not sending the amplitude parameter by DAQ I/F in Bias DAC is interpreted like a null demand by corresponding Bias DAC. A bad interpretation of the demand by Bias DAC is treated at item 2.2 Detection and switching (idem item 2.1)	3R	
2.4 Bis		BIAS	DF 2 : To transfer the controlling of specific amplitude to each Bias DAC	Failure of an output of Bias Logic: no transmission of an amplitude parameter relating to the signal of modulation of the spectrometric detectors	No generation of a numerical control relating to the spectrometric detectors towards Bias DAC (IF 1.2.2 degraded)	No sending towards the FPU of an analogue control of modulation signal towards the spectrometric detectors (degraded IF1.2)	Loss of all the data from the spectrometric detectors of the FPU towards the DPU (lost IF1) Need for switching on the redundant chain of the DRCU (except LIA) to recover all detectors data	DAQ I/F transfers to each Bias DAC via Bias Logic, the value of amplitude of the signal to be generated for the modulation of the spectrometric detectors. It is considered that not sending the amplitude parameter by DAQ I/F in Bias DAC is interpreted like a null demand by corresponding Bias DAC. A bad interpretation of the demand by Bias DAC is treated at item 2.2 Detection and switching (idem item 2.1)	3R	
2.5	DCU	BIAS	AF 4.2.1 : To transfer the analogue controls data towards the heaters of the JFET box (power supply of the JFET box)	Failure of photometric Heater DAC: no signal, setting OFF inopportune	No generation of analogue controls towards the JFET Box (no power supply of the heaters) (degraded IF1.2.2)	No heating of the JFET Box during the starting of a sequence of acquisition (degraded IF1.2)	Loss of all the data from the photometric detectors of the FPU towards the DPU (worst case) (lost IF1) Need for switching on the redundant chain of the DRCU (except LIA) to recover all detectors data	Detection of non heating by the Housekeeping parameters of SCU Each heater is individually commandable ON/OFF by means of low level commands	3R	
2.5 Bis	DCU	BIAS	AF 4.2.1 : To transfer the analogue controls data towards the heaters of the JFET box (power supply of the JFET box)	Failure of spectrometric Heater DAC: no signal, setting OFF inopportune	No generation of analogue controls towards the JFET Box (no power supply of the heaters) (degraded IF1.2.2)	No heating of the JFET Box during the starting of a sequence of acquisition (degraded IF1.2)	Loss of all the data from the spectrometric detectors of the FPU towards the DPU (worst case) (lost IF1) Need for switching on the redundant chain of the DRCU (except LIA) to recover all detectors data	Detection of non heating by the Housekeeping parameters of SCU Each heater is individually commandable ON/OFF by means of low level commands	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
2.6	DCU	BIAS	AF 4.2.1 : To transfer the analogue controls data towards the heaters of the JFET box (power supply of the JFET box)	Failure of photometric Heater DAC: permanent signal	Permanent generation of the analogue controls towards the JFET Box (permanent power supply of the heaters) (degraded IF1.2.2)	Permanent heating of the JFET Box during the sequences of acquisition (degraded IF1.2)	Loss of all the data from the photometric detectors of the FPU towards the DPU (worst case) (lost IF1) Risk of loss of all the data of the FPU towards DPU by thermal propagation (IF 1 lost) Switching on the redundant chain of the DRCU to recover all information detectors	An excessive temperature at the level of the JFET box is detected by the Housekeeping parameters of SCU Each heater bias must be commandable ON/OFF independantly by means of low level commands (control envisaged).	3R	
2.6 Bis	DCU	BIAS	AF 4.2.1 : To transfer the analogue controls data towards the heaters of the JFET box (power supply of the JFET box)	Failure of spectrometric Heater DAC: permanent signal	Permanent generation of the analogue controls towards the JFET Box (permanent power supply of the heaters) (degraded IF1.2.2)	Permanent heating of the JFET Box during the sequences of acquisition (degraded IF1.2)	Loss of all the data from the spectrometric detectors of the FPU towards the DPU (worst case) (lost IF1) Risk of loss of all the data of the FPU towards DPU by thermal propagation (IF 1 lost) Switching on the redundant chain of the DRCU to recover all information detectors	An excessive temperature at the level of the JFET box is detected by the Housekeeping parameters of SCU Each heater bias must be commandable ON/OFF independantly by means of low level commands (control envisaged).	3R	
2.7		BIAS	AF 4.2.3 : To ensure the transfer of continuous powering control of the JFET Box to allow the return of measurement information	Failure of a JFET DACs (no power supply, insufficient power supply, setting OFF inopportune)	No generation of continuous power supply or generation of an insufficient continuous power supply (IF 1.2.2 degraded)	No power supply or insufficient power supply of 24 channels of 2 JFET corresponding to the photometric detectors (IF 1.2 degraded)	Loss of 24 scientific data among a potential of 12*24 data (degraded IF1) switching on the redundant chain of the DRCU to recover all information detectors	Each return channel of a detector is composed of 2 JFET (amplifying of current) A JFET DAC (12 photometric) supplies 24 channels of detectors Detection of the loss of the data during the analysis by the scientists Each JFET power channel is individually commandable ON/OFF by means of low level commands (protection against the failures of a JFET box)	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
2.7 Bis		BIAS	AF 4.2.3 : To ensure the transfer of continuous powering control of the JFET Box to allow the return of measurement information	Failure of a JFET DACs (no power supply, insufficient power supply, setting OFF inopportune)	No generation of continuous power supply or generation of an insufficient continuous power supply (IF 1.2.2 degraded)	No power supply or insufficient power supply of 24 channels of 2 JFET corresponding to the spectrometric detectors (IF 1.2 degraded)	Loss of 24 scientific data among a potential of 3*24 data (degraded IF1) switching on the redundant chain of the DRCU to recover all information detectors	Each return channel of a detector is composed of 2 JFET (amplifying of current) A JFET DAC (3 spectrometric) supplies 24 channels of detectors Detection of the loss of the data during the analysis by the scientists Each JFET power channel is individually commandable ON/OFF by means of low level commands (protection against the failures of a JFET box)	3R	
2.8	DCU	BIAS	DF 4 : To transform and ensure the transfer of the sinusoidal signals towards Bias DACs	Failure of Sine DAC: No generation of an analogue signal Bad adaptation (amplitude, frequency, gain)	No analogue generation of control towards the BIAS DACs corresponding to the photometric detectors (lost AF 4.2.2; IF 1.2.2 degraded)	No sending of analogue control to the photometric detector of the FPU (IF 1.2 degraded)	Loss of all the data from the photometric detectors of the FPU towards the DPU (lost IF1) Need for switching on the redundant chain of the DRCU (except LIA) to recover all detectors data	Detection of the loss of the data during the analysis by the scientists The switching is carried out by the switching of power supply " Primary power line " (28 V) at the satellite level The switching procedure must be performed safely (shut off the blocks before switching of power supply).	3R	
2.8 Bis	DCU	BIAS	DF 4 : To transform and ensure the transfer of the sinusoidal signals towards Bias DACs	Failure of Sine DAC: No generation of an analogue signal Bad adaptation (amplitude, frequency, gain)	No analogue generation of control towards the BIAS DACs corresponding to the spectrometric detectors (lost AF 4.2.2; IF 1.2.2 degraded)	No sending of analogue control to the spectrometric detector of the FPU (IF 1.2 degraded)	Loss of all the data from the spectrometric detectors of the FPU towards the DPU (lost IF1) Need for switching on the redundant chain of the DRCU (except LIA) to recover all detectors data	Detection of the loss of the data during the analysis by the scientists The switching is carried out by the switching of power supply " Primary power line " (28 V) at the satellite level The switching procedure must be performed safely (shut off the blocks before switching of power supply).	3R	
2.9		BIAS	DF 4 : To transform and ensure the transfer of the sinusoidal signals towards Bias DACs	Heating of a Bias board (nominal or redundant) following internal failure	No generation of the numerical controls relating to the photometric detectors (IF 1.2.2 lost)	No sending towards the FPU of the analogue controls of modulation signal towards the photometric detectors (lost IF1.2)	Loss of all the data of the photometric detectors (IF 1 lost) Risk of propagation of failure (PoF) towards the other boards of the DRCU by thermal propagation (degradation of other DRCU boards). Total loss of the mission (loss of all the scientific data)	The Bias boards are redundant. The heating of a BIAS board is detected by a specific temperature measurement (Housekeeping parameters) returned towards the DPU. To avoid the thermal propagation towards other boards, it is necessary to insulate the board at fault quickly (cut off its supply via switch ON/OFF controlled by the SCU) and to switch on the redundant chain of the DRCU (except LIA) to recover all detectors data	2	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
2.9 Bis		BIAS	DF 4 : To transform and ensure the transfer of the sinusoidal signals towards Bias DACs	Heating of a Bias board (nominal or redundant) following internal failure	No generation of the numerical controls relating to the spectrometric detectors (IF 1.2.2 lost)	No sending towards the FPU of the analogue controls of modulation signal towards the spectrometric detectors (lost IF1.2)	Loss of all the data of the spectrometric detectors (IF 1 lost) Risk of propagation of failure (PoF) towards the other boards of the DRCU by thermal propagation (degradation of other DRCU boards). Total loss of the mission (loss of all the scientific data)	The Bias boards are redundant. The heating of a BIAS board is detected by a specific temperature measurement (Housekeeping parameters) returned towards the DPU. To avoid the thermal propagation towards other boards, it is necessary to insulate the board at fault quickly (cut off its supply via switch ON/OFF controlled by the SCU) and to switch on the redundant chain of the DRCU (except LIA) to recover all detectors data	2	
3.1		DAQ	AF 3.1.1 : To collect the numerical data of control coming from DPU	Failure of the DAQ - I/F (electronic or connector failure)	No reception of numerical information coming from DPU (IF 1.2.1 lost)	No generation of numerical controls relating to photometric detections towards the corresponding BIAS (Sine DAC, Heater DACs and JFET DACs) (IF 1.2 lost) No reception of the treated data (digitalisation) relating to the photometric detectors (IF 1.1 lost)	Loss of the 288 photometric scientific data of FPU towards DPU (IF 1 lost) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission	Idem item 2.1	2R	
3.1 Bis		DAQ	AF 3.1.1 : To collect the numerical data of control coming from DPU	Failure of the DAQ - I/F (electronic or connector failure)	No reception of numerical information coming from DPU (IF 1.2.1 lost)	No generation of numerical controls relating to spectrometric detections towards the corresponding BIAS (Sine DAC, Heater DACs and JFET DACs) (IF 1.2 lost) No reception of the treated data (digitalisation) relating to the spectrometric detectors (IF 1.1 lost)	Loss of the 72 spectrometric scientific data of FPU towards DPU (IF 1 lost) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission	Idem item 2.1 Bis	2R	
3.2	DCU	DAQ	AF 3.1.2 : To transfer the numerical data towards the BIAS	Failure of the DAQ - I/F (electronic or connector failure)	No generation of numerical controls towards Bias Logic of BIAS module (Sine DAC, Heater DACs and JFET DACs) (IF 1.2.1 lost)	No sending towards the FPU of the analogue modulation controls towards the detectors, heating controls of the JFET Box and power supply controls of the JFET (IF 1.2 lost)	Loss of the 288 photometric scientific data of FPU towards DPU (IF 1 lost) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission	Idem item 2.1	2R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
3.2 bis	DCU	DAQ	AF 3.1.2 : To transfer the numerical data towards the BIAS	Failure of the DAQ - I/F (electronic or connector failure)	No generation of numerical controls towards Bias Logic of BIAS module (Sine DAC, Heater DACs and JFET DACs) (IF 1.2.1 lost)	No sending towards the FPU of the analogue modulation controls towards the detectors, heating controls of the JFET Box and power supply controls of the JFET (IF 1.2 lost)	Loss of the 72 spectrometric scientific data of FPU towards DPU (IF 1 lost) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission	Idem item 2.1	2R	
3.3		DAQ	AF 3.1.2 : To transfer the numerical data towards the BIAS	Failure of DAQ I/F: bad amplitude parameter sent towards Bias DAC	Generation of an erroneous amplitude demand towards photometric Bias DAC (IF 1.2.1 degraded)	Transmission of a signal not adapted to a photometric detector of the FPU (degraded IF1.2)	Loss of all the data from the photometric detectors of the FPU towards the DPU (lost IF1) Risk of deterioration of the corresponding detectors (worst case; not compatible amplitude) Need for switching on the redundant chain of the DRCU (except LIA) to recover all detectors data	Amplitude parameters sent by the DAQ - I/F to each Bias Dac are returned to the DPU (Housekeeping parameters) following a DPU request . It should be noted that according to the failure mode (failure of output stages of DAQ I/F), the sending of a null amplitude demand by the DPU towards DAQ I/F is without effect	3R	
3.3 bis		DAQ	AF 3.1.2 : To transfer the numerical data towards the BIAS	Failure of DAQ I/F: bad amplitude parameter sent towards Bias DAC	Generation of an erroneous amplitude demand towards spectrometric Bias DAC (IF 1.2.1 degraded)	Transmission of a signal not adapted to a spectrometric detector of the FPU (degraded IF1.2)	Loss of all the data from the spectrometric detectors of the FPU towards the DPU (lost IF1) Risk of deterioration of the corresponding detectors (worst case; not compatible amplitude) Need for switching on the redundant chain of the DRCU (except LIA) to recover all detectors data	Amplitude parameters sent by the DAQ - I/F to each Bias Dac are returned to the DPU (Housekeeping parameters) following a DPU request . It should be noted that according to the failure mode (failure of output stages of DAQ I/F), the sending of a null amplitude demand by the DPU towards DAQ I/F is without effect	3R	
3.4		DAQ	AF 3.1.3 : To ensure the demodulation of analogue information at the LIA modules level	Loss or degradation of one of the 4 demodulation signals (failure of the DAQ - I/F; electronic or connector failure)	No sending of the demodulation signal towards LIA for the photometric detectors (IF 1.1.2 degraded)	No demodulation of the data received from the FPU relating to the photometric detectors (IF 1.1 degraded)	Loss of the data from the photometric detectors of the FPU towards the DPU (IF 1 degraded) Need for switching on the redundant chain of the DRCU (except LIA) to recover all detectors data	Idem item 2.1	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
3.4 Bis		DAQ	AF 3.1.3 : To ensure the demodulation of analogue information at the LIA modules level	Loss or degradation of one of the 2 demodulation signals (failure of the DAQ - I/F: electronic or connector failure)	No sending of the demodulation signal towards LIA for the spectrometric detectors (IF 1.1.2 degraded)	No demodulation of the data received from the FPU relating to the spectrometric detectors (IF 1.1 degraded)	Loss of the data from the spectrometric detectors of the FPU towards the DPU (IF 1 degraded) Need for switching on the redundant chain of the DRCU (except LIA) to recover all detectors data	Idem item 2.1	3R	
3.5		DAQ	AF 3.2.1 : To transfer the digitized data towards DPU	Failure of the DAQ - I/F (electronic or connector failure)	No numerical data processing for transfer to the DPU (IF 1.1.2 lost)	No transfer of the numerical data towards the DPU (IF 1.1 lost)	Loss of the 288 photometric scientific data from FPU towards DPU (IF 1 lost) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission	Idem item 2.1	2R	
3.5 Bis		DAQ	AF 3.2.1 : To transfer the digitized data towards DPU	Failure of the DAQ - I/F (electronic or connector failure)	No numerical data processing for transfer to the DPU (IF 1.1.2 lost)	No transfer of the numerical data towards the DPU (IF 1.1 lost)	Loss of the 72 spectrometric scientific data from FPU towards DPU (IF 1 lost) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission	Idem item 2.1	2R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
3.6	DCU	DAQ	AF 3.2.2 : To collect the analogue data demodulated, filtered and multiplexed by LIA	Failure of a multiplexer 4 > 1 (among 6)	Loss of 3 analogue channels of 16 photometric data (IF 1.1.2 degraded)	No transfer of the numerical data relating to 3 analogue channels of data among 18 (IF 1.1 degraded)	Loss of 48 photometric data among 288 from FPU towards DPU (IF 1 degraded) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission	6 multiplexers of 4 analogue channels of 16 data each (9 channels for P250, 6 for P350, 3 for P500, 4 for SW and 2 for LW; i.e. a total of 24 channels). Spectrometric channels are distributed on the 6 multiplexers. According to the distribution of the photometric analogue channels in input of the multiplexer, the loss of a multiplexer can lead to the loss of all the data of a type of detector (ex: loss of all the data of P500, if the 3 P500 channels are on the same multiplexer). A judicious distribution of the analogue channels could make possible to preserve data of each type of detector without switching on the redundant chain. The distribution will be carried out according to the priorities fixed by the scientists. If the 3 chains of T/C measuring (temperature) are used for the regulation of bolometers (via control of the cooler) and if each chain is segregated at the analogue channels level (see item 1.2), the distribution of the 3 analogue channels must be such as the loss of a multiplexer 4 > 1 does not lead to the loss of the 3 chains of temperature.	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
3.6 Bis	DCU	DAQ	AF 3.2.2 : To collect the analogue data demodulated, filtered and multiplexed by LIA	Failure of a multiplexer 4 > 1 (among 6)	Loss of an analogue channel of 16 spectrometric data (IF 1.1.2 degraded)	No transfer of the numerical data relating to 1 analogue channel of data among 6 (IF 1.1 degraded)	Loss of 16 spectrometric data among 72 from FPU towards DPU (IF 1 degraded) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission	6 multiplexers of 4 analogue channels of 16 data each (9 channels for P250, 6 for P350, 3 for P500, 4 for SW and 2 for LW; i.e. a total of 24 channels). Spectrometric channels are distributed on the 6 multiplexers. According to the distribution of the photometric analogue channels in input of the multiplexer, the loss of a multiplexer can lead to the loss of all the data of a type of detector (ex: loss of all the data of P500, if the 3 P500 channels are on the same multiplexer). A judicious distribution of the analogue channels could make possible to preserve data of each type of detector without switching on the redundant chain. The distribution will be carried out according to the priorities fixed by the scientists. If the 3 chains of T/C measuring (temperature) are used for the regulation of bolometers (via control of the cooler) and if each chain is segregated at the analogue channels level (see item 1.2), the distribution of the 3 analogue channels must be such as the loss of a multiplexer 4 > 1 does not lead to the loss of the 3 chains of temperature.	3R	
3.7	DCU	DAQ	AF 3.2.2 : To collect the analogue data demodulated, filtered and multiplexed by LIA	Failure of an OFFSET of an analogue channel among 6	No reduction of the noise relating to the 3 analogue channels of 16 photometric data Digitalisation of an analogue signal " disturbed " by the coder after amplification (IF 1.1.2 degraded)	Transfer of degraded numerical data towards the DPU (IF 1.1 degraded)	Sending of degraded data from FPU towards DPU: 48 photometric data among 288 (IF 1 degraded) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission	1 OFFSET by analogue / numerical coder	3R	
3.7 Bis	DCU	DAQ	AF 3.2.2 : To collect the analogue data demodulated, filtered and multiplexed by LIA	Failure of an OFFSET of an analogue channel among 6	No reduction of the noise relating to the 3 analogue channels of 16 spectrometric data Digitalisation of an analogue signal " disturbed " by the coder after amplification (IF 1.1.2 degraded)	Transfer of degraded numerical data towards the DPU (IF 1.1 degraded)	Sending of degraded data from FPU towards DPU: 16 spectrometric data among 72 (IF 1 degraded) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission	1 OFFSET by analogue / numerical coder	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
3.8		DAQ	AF 3.2.2 : To collect the analogue data demodulated, filtered and multiplexed by LIA	Failure of an OFFSET of an analogue channel among 6 : no transmission of the signal	Loss of 3 analogue channels of 16 photometric data (IF 1.1.2 degraded)	No transfer of the numerical data relating to 3 analogue channels of data among 18 (IF 1.1 degraded)	Loss of 48 photometric data among 288 from FPU towards DPU (IF 1 degraded) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission		3R	
3.8 Bis		DAQ	AF 3.2.2 : To collect the analogue data demodulated, filtered and multiplexed by LIA	Failure of an OFFSET of an analogue channel among 6 : no transmission of the signal	Loss of 3 analogue channels of an analogue channel of 16 spectrometric data (IF 1.1.2 degraded)	No transfer of the numerical data relating to 1 analogue channel of data among 6 (IF 1.1 degraded)	Loss of 16 spectrometric data among 72 from FPU towards DPU (IF 1 degraded) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission		3R	
3.9	DCU	DAQ	AF 3.2.2 : To collect the analogue data demodulated, filtered and multiplexed by LIA	Failure of an amplifier of an analogue channel among 6: bad amplification	Transfer for digitalisation of a " low " signal relating to the 3 analogue channels of 16 photometric data Digitalisation of an analogue signal badly amplified by the coder (IF 1.1.2 degraded)	Transfer of numerical data degraded towards the DPU (IF 1.1 degraded)	Sending of degraded data from FPU towards DPU: 48 photometric data among 288 (IF 1 degraded) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission	1 amplifier by analogue / numerical coder Non amplification of the signal leads to its non digitalisation (signal not in the dynamic range of the coder) and, consequently to the loss of the data (see item 3.10)	3R	
3.9 Bis	DCU	DAQ	AF 3.2.2 : To collect the analogue data demodulated, filtered and multiplexed by LIA	Failure of an amplifier of an analogue channel among 6: bad amplification	Transfer for digitalisation of a " low " signal relating to the analogue channel of 16 spectrometric data Digitalisation of an analogue signal badly amplified by the coder (IF 1.1.2 degraded)	Transfer of numerical data degraded towards the DPU (IF 1.1 degraded)	Sending of degraded data of FPU towards DPU: 16 spectrometric data among 72 (IF 1 degraded) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission	1 amplifier by analogue / numerical coder Non amplification of the signal leads to its non digitalisation (signal not in the dynamic range of the coder) and, consequently to the loss of the data (see item 3.10)	3R	
3.10.		DAQ	AF 3.2.2 : To collect the analogue data demodulated, filtered and multiplexed by LIA	Failure of an amplifier of an analogue channel among 6: no transmission of the signal	Loss of 3 analogue channels of 16 photometric data (IF 1.1.2 degraded)	No transfer of the numerical data relating to 3 analogue channels of data among 18 (IF 1.1 degraded)	Loss of 48 photometric data among 288 from FPU towards DPU (IF 1 degraded) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission		3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
3,10 Bis		DAQ	AF 3.2.2 : To collect the analogue data demodulated, filtered and multiplexed by LIA	Failure of an amplifier of an analogue channel among 6: no transmission of the signal	Loss of an analogue channel of 16 spectrometric data (IF 1.1.2 degraded)	No transfer of the numerical data relating to 1 analogue channel of data among 6 (IF 1.1 degraded)	Loss of 16 spectrometric data among 72 from FPU towards DPU (IF 1 degraded) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission		3R	
3,11	DCU	DAQ	AF 3.2.2 : To collect the analogue data demodulated, filtered and multiplexed by LIA	Failure of an analogue / numerical coder among 6 (no digitalisation of the data detector)	Loss of 3 analogue channels of 16 photometric data (IF 1.1.2 degraded)	No transfer of the numerical data relating to 3 analogue channels of data among 18 (IF 1.1 degraded)	Loss of 48 photometric data among 288 from FPU towards DPU (IF 1 degraded) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission		3R	
3,11Bis	DCU	DAQ	AF 3.2.2 : To collect the analogue data demodulated, filtered and multiplexed by LIA	Failure of an analogue / numerical coder among 6 (no digitalisation of the data detector)	Loss of an analogue channel of 16 spectrometric data (IF 1.1.2 degraded)	No transfer of the numerical data relating to 1 analogue channel of data among 6 (IF 1.1 degraded)	Loss of 16 spectrometric data among 72 from FPU towards DPU (IF 1 degraded) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission		3R	
3,12		DAQ	DF 1 : To transfer the numerical message outgoing from MTIM module towards DAQ I/F	Failure of an analogue / numerical coder (no transfer of the digitized data) or failure of a DAQ - I/F input	Loss of 3 analogue channels of 16 photometric data (IF 1.1.2 degraded)	No transfer of the numerical data relating to 3 analogue channels of data among 18 (IF 1.1 degraded)	Loss of 48 photometric data among 288 from FPU towards DPU (IF 1 degraded) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission	At the level of the DAQ - I/F, the 6 inputs relating to the 6 A/D converters are distinct	3R	
3,12 Bis		DAQ	DF 1 : To transfer the numerical message outgoing from MTIM module towards DAQ I/F	Failure of an analogue / numerical coder (no transfer of the digitized data) or failure of a DAQ - I/F input	Loss of an analogue channel of 16 spectrometric data (IF 1.1.2 degraded)	No transfer of the numerical data relating to 1 analogue channel of data among 6 (IF 1.1 degraded)	Loss of 16 spectrometric data among 72 from FPU towards DPU (IF 1 degraded) Require switching on the redundant chain of the DRCU (except LIA) to recover the mission	At the level of the DAQ - I/F, the 6 inputs relating to the 6 A/D converters are distinct	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
4.1	PSU	PSU	AF1.1 : To collect the electric power coming from S/C	Failure of a protection of the power supply bar, feeder or primary of the converter (open circuit)	Loss of power supply of a DC/DC converter of equipment (LIA-S, LIA-P, BIA-S, BIA-P, DAQ or SCU)	Loss of power supply of a module of the DCU (LIA, BIA, DAQ) (IF 1.1 and/or IF 1.2 lost) Loss of power supply of SCU module (IF 2.1, IF 2.2 and/or IF 2.3 lost)	Loss of all the scientific data issued from a type of detector of the FPU towards the DPU (lost IF1) Loss of the control functions (Housekeeping monitorings, calibration...) of DPU towards FPU (IF 2 lost) Need for switching on the redundant chain of the DRCU to recover the lost functions	Each PSU DC/DC converter shall have an independant electrical interface with the satellite power bus The short circuit located upstream of the secondary leads to the release of the protection of the power supply bar (module S/C) The loss of a secondary voltage is detected by the SCU (Housekeeping parameters) The switching is carried out by the switching of power supply " Primary power line " (28V) at the satellite level.	2R	
4.2		PSU	AF 1.2 : To transfer the electric power for distribution to the modules of the DRCU	Failure of the secondary of the transformer (open circuit)	Loss of power supply of equipment (LIA-S, LIA-P, BIA-S, BIA-P, DAQ or SCU)	Idem item 4.1	Idem item 4.1	Idem item 4.1	2R	
4.3	PSU	PSU	AF 1.2 : To transfer the electric power for distribution to the modules of the DRCU	Failure of the secondary of the transformer (short-circuit)	Loss of power supply of equipment (LIA-S, LIA-P, BIA-S, BIA-P, DAQ or SCU)	Idem item 4.1	Idem item 4.1	The power supply bar must be protected from short-circuit of a converter. This protection will be located as upstream as possible (preferably in specific equipment near to the generation / regulation / storage part).	2R	
4.4	PSU	PSU	AF 1.2 : To transfer the electric power for distribution to the modules of the DRCU	Overvoltage at output of the converter (short circuit of coils of the secondary)	Degradation of equipment following an overvoltage at output of converter	Deterioration of equipment of DCU module (LIA, BIA, DAQ) (IF 1.1 and/or IF 1.2 lost) Deterioration of equipment of SCU module (IF 2.1, IF 2.2 and/or IF 2.3 lost)	Loss of all the scientific data issued from a type of detector of the FPU towards the DPU (lost IF1) Loss of the control functions (Housekeeping monitorings, calibration...) from DPU towards FPU (IF 2 lost) Need for switching on the redundant chain of the DRCU to recover the lost functions The deterioration of the LIA blocks (9 photometric or 3 spectrometric) leads to the final loss of the corresponding data.	An overvoltage at output of the converter is detected by the secondary voltages (Housekeeping parameters of SCU). The detection of the secondary voltages could be located upstream of ON/OFF powering relays in order to guarantee the nominal voltage before the setting ON of the equipment. This implies that the DPU requests for the interrogation of the secondary voltages before controlling (via the SCU) the powering relays of the equipment. There is an overvoltage detection corresponding to nominal +10 %.	2R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
4.5	PSU	PSU	AF 1.2 : To transfer the electric power for distribution to the modules of the DRCU	Too important inrush current at the starting of equipment	/	/	Risk of collapse of the power supply bar of the satellite (PoF)	To be envisaged: a limiting device of inrush current in input of the converter However a progressive powering is considered. Moreover, the position of ON/OFF powering relays must be guaranteed before the powering of the "Primary power line (28V)" at the satellite level; i.e. the powering relays of the equipment must be in position OFF before their powering (see items related to function DF 4).	TBD	
5.1	SCU	SCU	AF 2.1 : To take delivery of information of control coming from DPU	Failure of the SCU Control Logic (failure of the input stage)	No reception of numerical information coming from DPU No generation of controls towards the support functions	No generation of numerical control relative: with the calibration (IF 2.2 lost) with the cooler (IF 2.3 lost) No generation of analogue control relative: with the temperatures of equipment FPU (IF 2.1 lost) with the secondary voltage and SCU/PSU temperature probes (lost AF 9) with switching ON/OFF of the equipment	Loss of the control functions (housekeeping, calibration, cooler...) from DPU towards FPU Loss of monitoring information of the equipment from FPU towards DPU (IF 2 lost) Need for switching on the redundant chain of the DRCU to recover the control and monitoring information	The failure of a SCU control logic is detected by the loss of all information at the DPU level: calibration, cooler, secondary voltages, housekeeping parameters. The switching on the redundant chain is carried out by the switching of power supply " Primary power line " (28 V) at the satellite level .	3R	
5.2	SCU	SCU	AF 2.2 : To ensure the return of information in answer to a DPU request	Failure of the SCU Control Logic (failure of the output stage towards DPU)	No transfer of monitoring information from the FPU, PSU and internal information of SCU, towards DPU	No sending of monitoring information relative to: the temperatures of equipment FPU (IF 2.1 lost) the calibration (IF 2.2 lost) the cooler (IF 2.3 lost) the secondary voltage and SCU/PSU temperature probes (lost AF 9) towards the DPU	Loss of monitoring information of the equipment from FPU towards DPU (IF 2 lost) Loss of Housekeeping information towards DPU (lost AF 9) Need for switching on the redundant chain of the DRCU to recover monitoring information	The failure of a SCU control logic is detected by the loss of all information at the DPU level: calibration, cooler, secondary voltages, housekeeping parameters.	3R	
5.3	SCU	SCU	AF 5.1 : To make carry out a control of temperature monitoring at the FPU level and ensure the return of corresponding information	Failure of T°C probe electronic module (intrinsic failure, inopportune switching off by a low level command, no power supply of probes, open circuit on a current limitation resistor)	No reception of one or more temperature measurements (among 16)	No sending of temperature information towards DPU (IF 2.1 lost)	Loss of the temperature monitoring information of the equipment from FPU towards DPU (worst case) (IF 2 lost) Need for switching on the redundant chain of the DRCU to recover temperature information	The controls of temperature monitoring are systematic. The temperature sensors are supplied by the SCU harness. Bias currents are individually and by a low level command switched off. The loss of temperature monitoring information is detected by the DPU.	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
5.4	SCU	SCU	AF 5.1 : To make carry out a control of temperature monitoring at the FPU level and ensure the return of corresponding information	Failure of T°C probe electronic module : too important bias current following short circuit or reduction in the resistivity of resistor dedicated to the limitation of current	Transmission of a signal not adapted towards a temperature sensor	Transmission of a signal not adapted towards a temperature sensor (IF 2.1 lost)	Loss of a temperature monitoring data of the equipment from FPU towards DPU (worst case) (IF 2 lost) Risk of deterioration of the corresponding temperature sensor (worst case: bias current not compatible) Risk of self-heating of probe (thermal dissipation), and consequently heating of the other equipment of FPU. See whether it requires to switch on the redundant chain of the DRCU.	Bias currents are individually and by a low level command switched off . A short-circuit at the level of a probe is protected by a current limitation resistor. A reduction in the resistivity is detected a priori by the DPU (non-coherent value of temperature) and must result in insulating the chain. The rise in temperature of the other equipment of FPU is detected by the DPU, and must lead to a period of setting off the DRCU, so that the equipment go down again in temperature. It is necessary to define the prerogatives of switching on the redundant DRCU according to the loss of only one temperature probe.	2	
5.5		SCU	DF 9 : Transfer of information between the Temperature control electronic module and the SCU control logic	Failure of T°C probe electronic module (intrinsic failure, failure of output stages towards DPU)	No sending of information of temperature towards the multiplexer 25 > 1	No sending of information of temperature towards DPU (IF 2.1 lost)	Loss of a temperature monitoring data of the equipment from FPU towards DPU (worst case) (IF 2 lost) Need for switching on the redundant chain of the DRCU to recover information of temperature	The loss of temperature monitoring information is detected by the DPU.	3R	
5.6	SCU	SCU	DF 9 : Transfer of information between the Temperature control electronic module and the SCU control logic	Failure of the multiplexer 25>1	No sending of information related to temperature, calibration and cooler, towards thermistor A/D for digitalisation (lost DF 7 and DF 8)	No sending of information related to temperature, calibration and cooler, towards DPU (IF 2.1, IF 2.2, IF 2.3 lost)	Loss of monitoring information of the equipment from FPU towards DPU (IF 2 lost) Need for switching on the redundant chain of the DRCU to recover monitoring information.	The loss of temperature monitoring information is detected by the DPU.	3R	
5.7		SCU	DF 9 : Transfer of information between the Temperature control electronic module and the SCU control logic	Failure of the amplifier located downstream from the multiplexer 25 > 1: bad amplification	Transfer for digitalisation by thermistors A/D of a " low " signal relating to the 25 analogue channels (16 temperature, 6 calibration and 3 cooler) (degraded DF 7 and DF 8) Digitalisation of an analogue signal badly amplified by the coder	Transfer of degraded numerical monitoring data towards DPU (IF 2.1, IF 2.2, IF 2.3 degraded)	Sending of degraded monitoring information from FPU towards DPU (degraded IF2) Need for switching on the redundant chain of the DRCU to recover monitoring information.	The degraded reception of monitoring information is detected by the DPU.	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
5.8		SCU	DF 9 : Transfer of information between the Temperature control electronic module and the SCU control logic	Failure of the amplifier located downstream from the multiplexer 25 > 1: no transmission of the signal	No sending of information related to temperature, calibration and cooler, towards thermistor A/D for digitalisation (lost DF 7 and DF 8)	No sending of information related to temperature, calibration and cooler, towards DPU (IF 2.1, IF 2.2, IF 2.3 lost)	Loss of monitoring information of the equipment from FPU towards DPU (IF 2 lost) Need for switching on the redundant chain of the DRCU to recover monitoring information.	The loss of monitoring information is detected by the DPU. Non amplification of the signal leads to its non digitalisation (signal out of the dynamic range of the coder), and consequently to the loss of the data (see item 5.9)	3R	
5.9	SCU	SCU	DF 9 : Transfer of information between the Temperature control electronic module and the SCU control logic	Failure of the N/A converter of thermistors AD (no digitalisation of information related to temperature, calibration and cooler, or no transfer to the SCU Control Logic) Failure of the SCU Control Logic (failure of the input stage associated with thermistors A/D)	Loss of information related to the monitoring of temperatures, calibration and cooler (lost DF 7 and DF 8)	No sending of information related to temperature, calibration and cooler, towards DPU (IF 2.1, IF 2.2, IF 2.3 lost)	Loss of monitoring information of the equipment from FPU towards DPU (IF 2 lost) Need for switching on the redundant chain of the DRCU to recover monitoring information.	The loss of monitoring information is detected by the DPU. The failures of the SCU Control Logic components in interface with the DPU are analysed in items 5.1 and 5.2	3R	
5.10.		SCU	DF 12 : To transmit the setting parameters (sampling rate) of the SCU control logic for the reception of information related to temperature, calibration and cooler	Failure of the Acquisition Sequencer module (no transmission of the sample rate)	No sending of information related to temperature, calibration and cooler towards the multiplexer 25 > 1	No sending of information related to temperature, calibration and cooler, towards DPU (IF 2.1, IF 2.2, IF 2.3 lost)	Loss of monitoring information of the equipment from FPU towards DPU (IF 2 lost) Need for switching on the redundant chain of the DRCU to recover monitoring information.	The packages of measurements are sent to the SCU according to the sample rate following a DPU request. All the acquisition chains are sampled at the same rate.	3R	
5.11	SCU	SCU	DF 12 : To transmit the setting parameters (sampling rate) of the SCU control logic for the reception of information related to temperature, calibration and cooler	Failure of the Acquisition Sequencer module (erroneous sample rate)	Sending of information related to temperature, calibration and cooler towards the multiplexer 25 > 1 with an erroneous sample rate	Sending of information related to temperature, calibration and cooler towards DPU with a sample rate different from that expected (IF 2.1, IF 2.2, IF 2.3 degraded)	Risk of use of erroneous information of equipment monitoring, from FPU towards DPU (IF 2 degraded) Risk of late detection of an anomaly on equipment of the FPU (if sample rate lower than that specified) (IF 2 degraded)	The sample rate is sent by the SCU following a DPU request. All the chains of acquisition are sampled at the same rate. The DPU can acquire information with a sample rate not corresponding to its request, but it cannot diagnose the anomaly. The DPU detects the loss of expected information. In this case, it is necessary to switch on the redundant chain of the DRCU to recover all information. A too high acquisition frequency which could lead to a risk of overflow, can disturb the treatment of the tasks and their realization by the DPU.	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
5,12	SCU	SCU	AF 5.2 : To make carry out a control of calibration of the detectors at the FPU level and ensure the return of corresponding information	Failure of a DAC (among 3) of the calibration module (no generation of an analogue signal)	No generation of control towards a calibration heater (1 photometric and 2 spectrometric)	No generation of analogue control relating to a calibration heater (IF 2.2 degraded)	Loss of the controls and the monitoring information of a calibration heater from DPU towards FPU (IF 2 degraded) Need for switching on the redundant chain of the DRCU to recover information of calibration Degradation of the spectrometric mission because of lack of reference (PoF)	For photometer calibrator, bias power is controlled by a low level command. For spectrometer calibrators, bias current are controlled by means of a low level command. The controls of calibration monitoring are systematic. Each DAC receives a demand of numerical current issued from the SCU following a DPU request. The loss of information (U and I) of calibration monitoring is detected by the DPU.	2R	
5,13	SCU	SCU	AF 5.2 : To make carry out a control of calibration of the detectors at the FPU level and ensure the return of corresponding information	Failure of a DAC (among 3) of the calibration module (bad adaptation, bad interpretation of the demand)	Sending of a current not adapted towards the corresponding amplifier	Sending of a current not adapted to a heater calibrator of the FPU (IF 2.2 degraded)	Loss of the controls and the monitoring information of a calibration heater from DPU towards FPU (IF 2 degraded) Need for switching on the redundant chain of the DRCU to recover information of calibration	The sending of a current not adapted towards a heater calibrator is detected by the DPU. The value of analogue current is retransmitted to the DPU (consistency check between the demand and the reread value). If a bad adaptation of the non compatible current of a heater leads to its deterioration, the rereading of the the current value must lead to the sending of a null current demand by the DPU. There is a limitation of current on the calibrators.	3R	
5,14	SCU	SCU	AF 5.2 : To make carry out a control of calibration of the detectors at the FPU level and ensure the return of corresponding information	Failure of an AOP or a DAC (among 3) of the calibration module (sending of a constant demand of setting ON of the Photometric calibrator)	Permanent control ON on the output of the calibration control module	Sending of a uninterrupted current on a heater calibrator of the FPU (IF 2,2 degraded)	Permanent calibration of the photometric detectors. Impossible measurement (IF 1 Lost)	Loss of the photometric mission because the detectors are masked by the source of calibration (PoF). Need for envisaging a control of setting OFF the source of calibration.	2R	
5,14 Bis	SCU	SCU	AF 5.2 : To make carry out a control of calibration of the detectors at the FPU level and ensure the return of corresponding information	Failure of an AOP or a DAC (among 3) of the calibration module (no generation of an analogue signal to the Spectrometric calibrator)	Permanent control OFF on the output of the calibration control module	No generation of an analogue control relating to a calibration heater (IF 2.2 degraded)	Slow degradation of the calibration of the Spectrometric detectors. (IF 1 degraded)		3R	
5,15	SCU	SCU	AF 5.2 : To make carry out a control of calibration of the detectors at the FPU level and ensure the return of corresponding information	Failure of an AOP (among 3) of the calibration module (No generation of an analogue signal)	Idem item 5.12	Idem item 5.12	Idem item 5.12	Idem item 5.12	2R	
5,16	SCU	SCU	AF 5.2 : To make carry out a control of calibration of the detectors at the FPU level and ensure the return of corresponding information	Failure of an AOP (among 3) of the calibration module (bad adaptation, bad interpretation of the demand)	Idem item 5.13	Idem item 5.13	Idem item 5.13	Idem item 5.13	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
5.17		SCU	AF 5.2 : To make carry out a control of calibration of the detectors at the FPU level and ensure the return of corresponding information	Failure of a DAC/AOP (among 3) of the calibration module (failure of the input stage in interface with FPU)	No reception of the tension and current measurement of a heater calibrator	No sending of information of a heater calibrator towards DPU (IF 2.2 degraded)	Loss of 2 monitoring information of a heater calibrator (IF 2 degraded) Need for switching on the redundant chain of the DRCU to recover the 2 information of a heater calibrator		3R	
5.18	SCU	SCU	DF 8 : Transfer of information between calibration monitoring and SCU Control Logic module	Failure of the calibration control module (intrinsic failure, failure of the output stage in interface with DPU)	No sending of calibration information towards the multiplexer 25 > 1	No sending of calibration information towards DPU (IF 2.2 lost)	Loss of monitoring information of the FPU calibration towards DPU (IF 2 lost) Need for switching on the redundant chain of the DRCU to recover information of temperature	The loss of calibration information is detected by the DPU.	3R	
5.19		SCU	DF 8 : Transfer of information between calibration monitoring and SCU Control Logic module	Failure of the multiplexer 25>1	Idem item 5.6 (DF 9 also lost)	Idem item 5.6	Idem item 5.6	Idem item 5.6	3R	
5.20		SCU	DF 8 : Transfer of information between calibration monitoring and SCU Control Logic module	Failure of the amplifier located downstream from the multiplexer 25 > 1: bad amplification	Idem item 5.7 (DF 9 also lost)	Idem item 5.7	Idem item 5.7	Idem item 5.7	3R	
5.21		SCU	DF 8 : Transfer of information between calibration monitoring and SCU Control Logic module	Failure of the amplifier located downstream from the multiplexer 25 > 1: no transmission of the signal	Idem item 5.8 (DF 9 also lost)	Idem item 5.8	Idem item 5.8	Idem item 5.8	3R	
5.22	SCU	SCU	DF 8 : Transfer of information between calibration monitoring and SCU Control Logic module	Failure of the N/A converter of thermistors AD (no digitalisation of information related to temperature, calibration and cooler, or no transfer to the SCU Control Logic) Failure of the SCU Control Logic (failure of the input stage associated with thermistors A/D)	Idem item 5.9 (DF 9 also lost)	Idem item 5.9	Idem item 5.9	Idem item 5.9	3R	
5.23	SCU	SCU	AF 5.3 : To make carry out a control on the level of the cooler and ensure the return of corresponding information	Failure of a DAC / AOP (among 3) of the cooler control module (no generation of an analogue signal)	No generation of a ON control towards HSE	No generation of an control to the cooler (control ON of HSE) (AF 5.3 lost)	Loss of the cryocooler functionality. Loss of all observation data. (IF 1 lost) Need for switching on the redundant chain of the DRCU to recover the control of the cooler	Redundancy of coolers. The controls of control of the cooler are systematic. Each DAC receives an demand of numerical current issued from the SCU following a request DPU. The loss of the monitoring voltage information of each heater of the cooler is detected by the DPU.	2R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
5.24	SCU	SCU	AF 5.3 : To make carry out an control on the level of the cooler and ensure the return of corresponding information	Failure of a DAC / AOP (among 3) of the cooler control module (no generation of an analogue signal)	No generation of a OFF control control towards HSE	No generation of an control to the cooler (control OFF of HSE) (AF 5.3 lost)	Recycling impossible to perform at the cooler level: reduction of the mission duration. Degradation of observation data. (IF 1 degraded) Need for switching on the redundant chain of the DRCU to recover the control of the cooler	The controls of control of the cooler are systematic. Each DAC receives an demand of numerical current issued from the SCU following a request DPU. The loss of the monitoring voltage information of each heater of the cooler is detected by the DPU.	3R	
5.25	SCU	SCU	AF 5.3 : To make carry out an control on the level of the cooler and ensure the return of corresponding information	Failure of a DAC / AOP (among 3) of the cooler control module (no generation of an analogue signal)	No generation of a ON control control towards HSP	No generation of an control to the cooler (control ON of HSP) (AF 5.3 lost)	Recycling impossible to perform at the cooler level: reduction of the mission duration. Degradation of observation data. (IF 1 degraded) Need for switching on the redundant chain of the DRCU to recover the control of the cooler	The controls of control of the cooler are systematic. Each DAC receives an demand of numerical current issued from the SCU following a request DPU. The loss of the monitoring voltage information of each heater of the cooler is detected by the DPU.	3R	
5.26	SCU	SCU	AF 5.3 : To make carry out an control on the level of the cooler and ensure the return of corresponding information	Failure of a DAC / AOP (among 3) of the cooler control module (no generation of an analogue signal)	No generation of a OFF control control towards HSP	No generation of an control to the cooler (control OFF of HSP) (AF 5.3 lost)	Loss of the cryocooler functionality. Loss of all observation data. (IF 1 lost) Need for switching on the redundant chain of the DRCU to recover the control of the cooler	Redundancy of coolers. The controls of control of the cooler are systematic. Each DAC receives an demand of numerical current issued from the SCU following a request DPU. The loss of the monitoring voltage information of each heater of the cooler is detected by the DPU.	2R	
5.27	SCU	SCU	AF 5.3 : To make carry out an control on the level of the cooler and ensure the return of corresponding information	Failure of a DAC / AOP (among 3) of the cooler control module (no generation of an analogue signal)	Sending of a current not adapted towards the corresponding HSE or HSP amplifier	Sending of a current not adapted on Heater cooler of the FPU (IF 2.2 degraded)	Loss of the control and the monitoring information of a heater of the cooler from DPU towards FPU (IF 2 degraded) Need for switching on the redundant chain of the DRCU to recover the control of the cooler	The sending of a current not adapted towards Heater of the cooler is detected by the DPU (consistency check between the demand of current and the reread voltage value). If a bad adaptation of the non compatible current of a heater leads to its deterioration, the rereading of the voltage value must lead to the sending of a null current demand by the DPU. There is an absence of variation in temperature of the sorption pump because a detection is included in the algorithm of the recycling control.	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
5.28	SCU	SCU	AF 5.3 : To make carry out a control at the level of the cooler and ensure the return of corresponding information	Failure of a DAC/AOP (among 3) of the cooler module (failure of the input stage in interface with FPU)	No reception of the voltage measurement of a cooler heater	No sending of the information of a cooler heater towards DPU (IF 2.3 degraded)	Loss of the monitoring information of a cooler heater (IF 2 degraded) Need for switching on the redundant chain of the DRCU to recover information of a cooler heater		3R	
5.29		SCU	DF 7 : Transfer of information between cooler control and the SCU control logic module	Failure of the cooler control module (intrinsic failure, failure of the output stage in interface with DPU)	No sending of cooler information towards the multiplexer 25 > 1	No sending of cooler information towards DPU (IF 2.3 lost)	Loss of monitoring information of the cooler from FPU towards DPU (IF 2 lost) Need for switching on the redundant chain of the DRCU to recover cooler information	The loss of cooler information is detected by the DPU.	3R	
5.30		SCU	DF 7 : Transfer of information between cooler control and the SCU control logic module	Failure of the multiplexer 25>1	Idem item 5.6 (DF 9 also lost)	Idem item 5.6	Idem item 5.6	Idem item 5.6	3R	
5.31		SCU	DF 7 : Transfer of information between cooler control and the SCU control logic module	Failure of the amplifier located downstream from the multiplexer 25 > 1: bad amplification	Idem item 5.7 (DF 9 also lost)	Idem item 5.7	Idem item 5.7	Idem item 5.7	3R	
5.32		SCU	DF 7 : Transfer of information between cooler control and the SCU control logic module	Failure of the amplifier located downstream from the multiplexer 25 > 1: no transmission of the signal	Idem item 5.8 (DF 9 also lost)	Idem item 5.8	Idem item 5.8	Idem item 5.8	3R	
5.33	SCU	SCU	DF 7 : Transfer of information between cooler control and the SCU control logic module	Failure of the N/A converter of thermistors AD (no digitalisation of information related to temperature, calibration and cooler, or no transfer to the SCU Control Logic) Failure of the SCU Control Logic (failure of the input stage associated with thermistors A/D)	Idem item 5.9 (DF 9 also lost)	Idem item 5.9	Idem item 5.9	Idem item 5.9	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
5.34		SCU	DF 10 : To ensure the transfer of secondary voltages information towards the HK / AD module	Failure of a multiplexer 18 > 1 (among 2)	No sending of 18 secondary voltages information towards the HK A/D module for digitalisation	No sending of 18 secondary voltages information towards DPU (degraded IF3)	Loss of 18 monitoring information of the PSU equipment towards FPU, among a potential of 36 (degraded AF 9) Risk of non detection of an overvoltage at the converter output, which can lead to the deterioration of equipment Risk of non detection of a loss of equipment supply	The controls of secondary voltages monitoring are performed on request. See item 4.4 for the risk of overvoltage at the converter output. The loss of equipment supply not detected by the secondary voltage information can be diagnosed by other parameters (ex: loss of power supply of a SCU board by non reception of information related to calibration, cooler and temperature by the DPU) .	3R	
5.35	SCU	SCU	DF 10 : To ensure the transfer of secondary voltages information towards the HK / AD module	Failure of the amplifier located downstream from the 2 multiplexers 18 > 1: bad amplification	Transfer for digitalisation by the HK A/D module of a low signal relating to 36 secondary voltages information	Transfer of degraded numerical information of secondary voltages towards DPU (degraded AF 9)	Sending of degraded information of equipment monitoring from PSU towards DPU (degraded AF 9) Need for switching on the redundant chain of the DRCU to recover monitoring information	The degraded reception of secondary voltages information is detected by the DPU following a request	3R	
5.36		SCU	DF 10 : To ensure the transfer of secondary voltages information towards the HK / AD module	Failure of the amplifier located downstream from the 2 multiplexers 18 > 1: no transmission of the signal	No sending of secondary voltages information towards the HK A/D module for digitalisation	No sending of secondary voltages information towards DPU (degraded IF3)	Loss of monitoring information of the PSU equipment towards FPU (degraded AF 9)	The loss of monitoring information is detected by the DPU following a request Non amplification of the signal leads to non digitalisation (signal is out of the dynamic range of the coder), and consequently to the loss of the data (see item 5.40)	3R	
5.37	SCU	SCU	DF 10 : To ensure the transfer of secondary voltages information towards the HK / AD module	Failure of the multiplexer 4>1	No sending of secondary voltages information towards the HK A/D module for digitalisation (AF 7 and DF 11 also lost)	No sending of information related to secondary voltages, SCU temperature and PSU temperature towards DPU (lost AF 9)	Loss of monitoring information related to the equipment of PSU and SCU towards FPU (lost AF 9) Risk of non detection of a rise in temperature of SCU equipment non compatible with the mission (thermal propagation which can lead to a degradation of the bolometric data)	The monitoring controls of secondary voltages and SCU temperature probes are performed on request. See item 5.34, 5.35 and 5.36 for secondary voltages information.	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
5.38		SCU	DF 10 : To ensure the transfer of secondary voltages information towards the HK / AD module	Failure of the amplifier located downstream from the multiplexer 4 > 1: bad amplification	Transfer for digitalisation by the HK A/D module of a low signal relating to 36 secondary voltages information (also lost AF 7 and DF 11)	Transfer of degraded numerical information related to secondary voltages, SCU temperature and PSU temperature, towards DPU (degraded AF 9)	Sending of degraded monitoring information of PSU and SCU equipment towards DPU (degraded AF 9) Need for switching on the redundant chain of the DRCU to recover monitoring information	The degraded reception of information related to secondary voltages and SCU temperatures is detected by the DPU following a request	3R	
5.39	SCU	SCU	DF 10 : To ensure the transfer of secondary voltages information towards the HK / AD module	Failure of the amplifier located downstream from the multiplexer 4 > 1: no transmission of the signal	No sending of secondary voltages information towards the HK A/D module for digitalisation (AF 7 and DF 11 also lost)	No sending of information related to secondary voltages, SCU temperature and PSU temperature towards DPU (lost AF 9)	Loss of monitoring information related to the equipment of PSU and SCU towards FPU (lost AF 9)	The loss of monitoring information is detected by the DPU following a request Non amplification of the signal leads to non digitalisation (signal is out of the dynamic range of the coder), and consequently to the loss of the data (see item 5.40)	3R	
5.40.		SCU	DF 10 : To ensure the transfer of secondary voltages information towards the HK / AD module	Failure of the numerical analogue converter (no digitalisation of the data)	Loss of information of secondary voltage and SCU temperature probes (AF 7 and DF 11 also lost)	No sending of information related to secondary voltages, SCU temperature and PSU temperature towards DPU (lost AF 9)	Loss of monitoring information related to the equipment of PSU and SCU towards FPU (lost AF 9)	The loss of information is detected by the DPU	3R	
5.41		SCU	DF 5 : To ensure the transfer of information from the HK A/D module towards the SCU Control Logic module	Failure of A/D converter (no transfer to the SCU Control Logic) Failure of SCU Control Logic (failure of the input stage associated with HK A/D)	Loss of information related to secondary voltages, SCU temperature and PSU temperature	No sending of information related to secondary voltages, SCU temperature and PSU temperature towards the DPU (lost AF 9)	Loss of monitoring information related to the equipment of PSU and SCU towards FPU (lost AF 9)	The loss of information is detected by the DPU The failures of the SCU control logic components in interface with the DPU are analysed at items 5.1 and 5.2	3R	
5.42		SCU	DF 13 : To transmit the setting parameters (frequency...) of the SCU Control Logic for the reception of information related to secondary voltages and SCU/PSU temperature probes	Failure of the Acquisition Sequencer module (no transmission of the sample rate)	No sending of information related to secondary voltages, SCU temperature and PSU temperature towards the multiplexer 4 > 1	No sending of information related to secondary voltages, SCU temperature and PSU temperature towards the DPU (lost AF 9)	Loss of monitoring information related to the equipment of PSU and SCU towards FPU (lost AF 9) Need for switching on the redundant chain of the DRCU to recover monitoring information.	The sample rate is sent by the SCU following a DPU request. All the acquisition chains are sampled at the same rate.	3R	
5.43	SCU	SCU	DF 13 : To transmit the setting parameters (frequency...) of the SCU Control Logic for the reception of information related to secondary voltages and SCU/PSU temperature probes	Failure of the Acquisition Sequencer module (erroneous sample rate)	Sending of information related to secondary voltages, SCU temperature and PSU temperature towards the multiplexer 4 > 1 with an erroneous sample rate	Sending of information related to secondary voltages, SCU temperature and PSU temperature towards the DPU with a sample rate different from expected (lost AF 9)	Risk of use of erroneous information related to PSU and SCU monitoring towards DPU (lost AF 9) Risk of late detection of an anomaly on PSU or SCU equipment (if sample rate lower than specified) (lost AF 9)	Idem item 5.11	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
5.44		SCU	AF 7 : To ensure the transfer of PSU temperature probes information towards the HK / AD module		See items 5.37 to 5.40	See items 5.37 to 5.40	See items 5.37 to 5.40	<p>The loss of monitoring information is detected by the DPU following a request</p> <p>The equipment supporting function AF 7 are identical to those associated with function DF 10 (multiplexer 4 > 1 and HK A/D module). The failures of these equipment are analysed at items 5.37 to 5.40</p> <p>The failures of the SCU control logic components in interface with the DPU are analysed at items 5.1 and 5.2</p> <p>The PSU is not part of the SCU. It is considered that temperature information of PSU module will be retransmitted as housekeeping parameters.</p>	3R	
5.45		SCU	DF 11 : To ensure the transfer of SCU temperature probes information towards the HK / AD module		See items 5.37 to 5.40	See items 5.37 to 5.40	See items 5.37 to 5.40	<p>The loss of monitoring information is detected by the DPU following a request</p> <p>The 3 temperatures at the SCU level correspond to the 3 boards which constitute it.</p> <p>The equipment supporting function DF 11 are identical to those associated with function DF 10 (multiplexer 4 > 1 and HK A/D module). The failures of these equipment are analysed at items 5.37 to 5.40.</p> <p>The failures of the SCU control logic components in interface with the DPU are analysed at items 5.1 and 5.2.</p>	3R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
5.46	SCU	PSD	AF 6 : To ensure the transfer of power supply towards DCU module	Failure of "external harness" connector (pin failure, open circuit...) Failure of the distribution line (open circuit) Galvanic insulation failure	No transfer of power supply towards an element of the DCU (LIA P, LIA S, BIAS, DAQ)	Loss of power supply of DCU equipment (BIAS, DAQ or LIA)	Loss of all the data from the detectors (photometric or spectrometric) (worst case: loss of power supply of LIA P or LIA S board) (IF 1 lost) The loss of power supply of a BIAS or DAQ must lead to the switching on the redundant chain DRCU to recover all detectors data.	The photometric and spectrometric modes detection are exclusive modes. In photometric mode (resp. spectrometric), the 3 LIA_S boards are not supplied (resp. the 9 LIA_P boards). The choice of the mode is carried out by the DPU. Distribution of power supply (+ 5.2 and? 9 V) of LIA is carried out by 2 relays in series. A ON/OFFrelay is controlled by the SCU following a DPU request (see function DF 6). The other relay is directly controlled by the choice of Primary Power Line (28 V) supply at the satellite level (see function AF 8). The supply of the DCU elements is carried out via an external harness. The failures related to the converters are analysed at items 4.2 to 4.5.	2R	
5.47	SCU	SCU	DF 6 :To configure the relays of power supply of the DRCU elements	No setting ON of a powering relay (coil burn out, failed contact, loss of 12 V " SCU_REL_P12 ",...) Inopportune setting OFF of a powering relay (short circuit on coil, stuck contact...)	No power supply or loss of supply of DRCU equipment	Loss of power supply of a module of the DCU (LIA, BIAS, DAQ) (IF 1.1 and / or IF 1.2 lost) Loss of power supply of SCU module (IF 2.1, IF 2.2 and/or IF2.3 lost)	Loss of all the scientific data issued from a type of detector (photometric or spectrometric) from FPU towards DPU (lost IF1) Loss of the monitoring functions (Housekeepink, calibration, cooler) from DPU towards FPU (IF 2 lost) Need for switching on the redundant chain of the DRCU to recover to the lost function(s).	The control of the powering relays ("bistable" type relay) is carried out by SCU following a DPU request. When DRCU is in state OFF, all the relays are in state OFF. The loss of power supply of a DCU or SCU module is detected by the DPU, by the loss of information relating to the equipment not supplied (loss of the scientific data if loss of DCU supply, and loss of the monitoring data if loss of SCU supply). Moreover, according to the implementation of secondary voltages detection (see item 4.4), the loss of power supply can be detected by the "Secondary voltages" Housekeeping parameters.	2R	
5.48	SCU	SCU	DF 6 :To configure the relays of power supply of the DRCU elements	Inopportune setting ON of a powering relay (short circuit on coil, stuck contact...)	Accidental power supply of an element of the DRCU		The DRCU consumes a great quantity of energy, so the satellite will cut off its supply. DRCU is "switched off".	Under nominal operation, the relays of powering of the elements of the chain are in position ON (except specific case of supply of LIA P or LIA S; see 5.46 item). See interface with the MCU	2	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
5.49	SCU	SCU	DF 6 :To configure the relays of power supply of the DRCU elements	No setting OFF of powering relay (coil burn out, failed contact, short circuit downstream, heating, loss of 12 V " SCU_REL_P12 "...)	No power supply cut-off of an element of the DRCU	No switching of detection mode (from photometric to spectrometric and vice versa) No insulation of the DRCU equipment following internal failure (overheating, short circuit...)	Loss of all the scientific data issued from a type of detector (photometric or spectrometric) (IF 1 lost) Need for switching on the redundant chain of the DRCU to recover the detectors data. Permanent consumption and risk of propagation of failure by thermal dissipation. Risk of total loss of the mission.	If possible, dimension the protections to make acceptable the thermal dissipation generated by any short-circuit downstream. Dimension the circuit of setting OFF (resistance to heating) and the response time of the monitoring with respect to the risk of propagation of failure by thermal dissipation. If necessary, make reliable the setting OFF of equipment that consumes the most. The photometric and spectrometric modes are exclusive (see item 5.46). If LIA P and LIA S are supplied simultaneously, a double control of DPU allows the setting ON of a type of LIA and the setting OFF the other type.	2R	
5.50	SCU	SCU	DF 6 :To configure the relays of power supply of the DRCU elements	Powering relays in the not supplied redundant chain of the DRCU remained in position ON [loss of power supply of SCU (SCU_DIG_P5) or loss of the power supply of the powering relays (SCU_REL_P12)]	Incoherent configuration of the powering relays in regard with the chain of the DRCU (not supplied chain)	Too important inrush current at the the starting of equipment following the switching on the corresponding chain (chain in which all the powering relays of are in position ON)	Risk of collapse of the satellite bar	See whether it is necessary that the SCU forces all the powering relays on state OFF at the time of their initialization (starting following the powering of the corresponding relay)	TBD	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
5.51	SCU	SCU	AF 8 : To configure the power supply M/R (Main / Redundant) of the LIA boards of DCU	Failure of the power on pulse generators or failure of a M/R configuration relay	Misconfiguration of the relay in regard with the power supply bar (relay on Main position whereas Redundant 28 V is present)	No power supply of LIA_P or LIA_S modules (IF 1.1 lost)	Loss of all the scientific data (photometric or spectrometric) from FPU towards DPU (lost IF1) Need for switching on the redundant chain of the DRCU to recover the lost function(s)	Distribution of power supply (+ 5.2 and 9 V) of LIA boards is carried out by 2 relays in series. A ON/OFF relay is controlled by the SCU following a DPU request (see function DF 8). The other relay is directly controlled by the choice of Primary Power Line (28 V) supply at the satellite level. This relay avoids to create a common point between the secondaries of power supply of redundant LIA and main LIA associated with spectrometers and photometers. This relay is positioned in the configuration M or R by means of a pulsed detector (power on pulse generators) implemented on the power supply line of 5 V SCU. The pulse is effective following the voltage presence at the level of the 5 V SCU (voltage obtained by the satellite power supply bar ; Main or Redundant). Piloting these relays requires the 12V dedicated to controls (SCU_REL_P12).	2R	
5.51	SCU	SCU	AF 8 : To configure the power supply M/R (Main / Redundant) of the LIA boards of DCU (cont'd - 1)					If the voltage rising of 12 V (SCU_REL_P12) for controlling the relays is slower than the voltage rising 5 V SCU (SCU_DIG_P5), and consequently of the pulse duration of the voltage presence detector, the relays of configuration M/R of the LIA boards will not move. To guarantee the position of relays M/R, it is necessary either that the pulse duration is compatible of the voltage rising of 12 V for relay control, or that voltage presence detectors are implemented (power pulse-on generators) on the line of 12 V for relays (SCU_REL_P12) (the pulse will be effective when 12 V is nominal).	2R	

Item	Module	Block	Functions	Failure modes	Local effects	Module effects	Subsystem effects (DRCU)	Remarks	Classification of the effects at DRCU level	Classification of the effects at instrument level
5.51	SCU	SCU	AF 8 : To configure the power supply M/R (Main / Redundant) of the LIA boards of DCU (cont'd - 2)					<p>It should be noted that the loss of a voltage presence detector (power pulse-on generators) leads to the loss of the possibility of switching on the redundant chain of the DRCU, and constitutes a dormant failure not detected before switching.</p> <p>See whether it is necessary to redate the voltage presence detectors. See whether it is necessary to set up a detection relating to the correct operation of these detectors. However it is necessary to find a compromise between the addition of equipment and the loss of reliability they can generate.</p>	2R	