

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 1 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

1.	INTRODUCTION	2
1.1	Change Record.....	2
1.2	Applicable Documents.....	2
1.3	Reference Documents	2
1.4	General instructions for executing test procedures	2
1.5	Assumptions.....	2
1.6	Open Issues	3
2.	COLD FUNCTIONAL TEST PROCEDURES	4
2.1	Procedure SPIRE-IST-DPU-ON.....	4
2.2	Procedure SPIRE-IST-DRCU-ON.....	5
2.3	Procedure SPIRE-IST-FUNC-SCU-02.....	6
2.4	Procedure SPIRE-IST-FUNC-SCU-03.....	7
2.5	Procedure SPIRE-IST-FUNC-SCU-06.....	8
2.6	Procedure SPIRE-IST-FUNC-SCU-07.....	9
2.7	Procedure SPIRE-IST-FUNC-SCU-04.....	10
2.8	Procedure SPIRE-IST-FUNC-PCAL-01	11
2.9	Procedure SPIRE-IST-FUNC-SCU-05.....	12
2.10	Procedure SPIRE-IST-FUNC-SCAL-01	13
2.11	Procedure SPIRE-IST-FUNC-MCU-01	14
2.12	Procedure: SPIRE-IST-FUNC-MCU-03.....	15
2.13	Procedure SPIRE-IST-FUNC-BSM-01.....	16
2.14	Procedure SPIRE-IST-FUNC-BSM-03.....	17
2.15	Procedure SPIRE-IST-FUNC-BSM-05a.....	18
2.16	Procedure SPIRE-IST-FUNC-BSM-05b.....	19
2.17	Procedure SPIRE-IST-BSM-OFF	20
2.18	Procedure SPIRE-IST-FUNC-SMEC-02a	21
2.19	Procedure SPIRE-IST-FUNC-SMEC-01	22
2.20	Procedure SPIRE-IST-FUNC-SMEC-03	23
2.21	Procedure SPIRE-IST-FUNC-SMEC-04a	24
2.22	Procedure SPIRE-IST-FUNC-SMEC-09	25
2.23	Procedure SPIRE-IST-FUNC-SMEC-04b	26
2.24	Procedure SPIRE-IST-FUNC-SMEC-07	27
2.25	Procedure SPIRE-IST-FUNC-SMEC-06	28
2.26	Procedure SPIRE-IST-SMEC-OFF.....	29
2.27	Procedure SPIRE-IST-FUNC-SMEC-02b	30
2.28	Procedure SPIRE-IST-FUNC-DCU-01.....	31
2.29	Procedure SPIRE-IST-FUNC-DCU-11-P	32
2.30	Procedure SPIRE-IST-FUNC-DCU-13-P	33
2.31	Procedure SPIRE-IST-PDET-OFF.....	34
2.32	Procedure SPIRE-IST-FUNC-DCU-11-S	35
2.33	Procedure SPIRE-IST-FUNC-DCU-13-S	36
2.34	Procedure SPIRE-IST-SDET-OFF.....	37
2.35	Procedure SPIRE-IST-MCU-OFF.....	38
2.36	Procedure SPIRE-IST-SCU-OFF.....	39
2.37	Procedure SPIRE-IST-DRCU-OFF.....	40
2.38	Procedure SPIRE-IST-DPU-OFF.....	41

1. INTRODUCTION

This document contains the SPIRE Cold Functional Test Procedures to be executed during the Herschel IST. These procedures, which are valid **for both He I and He II temperatures**, are listed here in the order in which they are expected to be performed. These procedures give step-by-step instructions on how to perform each test. Some procedures can only be run after integration – where appropriate this is clearly indicated in the preconditions section of each procedure.

1.1 Change Record

Issue 0.1, 27/04/2005 – Draft version for EQM

Issue 2.0, 12/06/2006 – First version for IST

1.2 Applicable Documents

AD01 SPIRE Functional Test Specification, Issue 1.4, SPIRE-RAL-DOC-001652, 22/07/2005

AD02 SPIRE ILT Cold Functional Test Procedure, Issue 1.1, SPIRE-RAL-PRC-002585, 22/03/2006

1.3 Reference Documents

RD01 SPIRE Instrument User Manual, Issue 1.0, SPIRE-RAL-PRJ-002395, 08/04/2005

1.4 General instructions for executing test procedures

- Before executing any of the procedures please always check with the I-EGSE staff
- Any text in **boldface** in the procedural steps generally indicates an action which has to be performed manually by the Instrument EGSE (I-EGSE) staff.
- The procedure tables include blank boxes where the actual values of parameters can be noted. Based on the comparison with the expected values the success or failure of a step should be recorded in the final column of the table.
- The last row in a procedure table should be used to record the overall Pass/Fail result of each test.

1.5 Assumptions

- The CCS are only required to check changes in instrument configuration related HK parameters.
- For these functional tests the instrument will not always be in a pre-defined mode as listed in the IUM (**RD01**).

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 3 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

- For the SPIRE spectrometer mechanism (SMECm) tests it is assumed that the Herschel cryostat will be tilted (TBD).
- **These procedures should be suitable for operation of both the Prime and Redundant side of the instrument (TBD).**
- **The converted TM parameter values are extracted from the MIB in use for PFM ILT. These values are subject to change for both prime and redundant operations.**
- **The SPIRE thermometry channel values quoted in the procedures are dependent on the Herschel cryostat achieving nominal Level 0 and Level 1 temperatures.**

1.6 Open Issues

- Procedures for dealing with instrument contingencies will be addressed in the next version of this document.
- Names of the Herschel Satellite procedures for powering on/off the SPIRE DPU and DRCU are to be filled in the next version. In this version they are marked as procedure XXXXXX.

1.7 Duration

The estimated duration for executing the entire CFT sequence of procedures, including switch off of the SPIRE instrument afterwards is estimated to be about **4 hours**.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 4 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2. COLD FUNCTIONAL TEST PROCEDURES

2.1 Procedure SPIRE-IST-DPU-ON

Version: 1.0

Date: 12th June 2006

Purpose: To switch on the SPIRE DPU and start generating housekeeping

Duration: 2 minutes plus time to execute CCS procedure XXXXXX

Preconditions:

- Procedure to supply 28V Power Supply from the satellite to the SPIRE DPU is available
- SPIRE MIB is imported in the CCS database.
- CCS is up and running (SCOS, TOPE and the CDMU)
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS
- The I-EGSE is up and running

Initial Configuration: SPIRE Warm Electronics (DPU and DRCU) are switched off

Procedure Steps:

Step	Description	Parameter	Expected Values Before/After	Actual Values Before/After	Pass/Fail
1	Using CCS procedure XXXXX Power on the SPIRE DPU 28V Power Supply	—	—	—	
2	Wait for instruction from I-EGSE staff to continue with the procedure	—	—	—	
3	Execute TCL script SPIRE-IST-DPU-ON.tcl	—	—	—	
4	Check that THSK parameter is refreshing every second	—	—	—	
5	Check that TM2N parameter is incrementing every second	—	—	—	

Test Result (Pass/Fail):

Final Configuration: SPIRE DPU is on but the DRCU is still off

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 5 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.2 Procedure SPIRE-IST-DRCU-ON

Version: 1.0

Date: 12th June 2006

Purpose: To switch on the SPIRE DRCU and start generating housekeeping

Duration: 4 minutes

Preconditions:

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched off
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Procedure Steps:

Step	Description	Parameter	Expected Values Before/After	Actual Values Before/After	Success/Failure
1	Execute TCL script SPIRE-IST-DRCU-ON-STEP1.tcl	—	—	—	
2	Check that THSK parameter is not refreshing anymore	—	—	—	
3	Check that TM2N parameter is not incrementing anymore	—	—	—	
4	When instructed by the I-EGSE staff Power on the SPIRE DRCU using the CCS procedure XXXXXX	—	—	—	
5	Execute TCL script SPIRE-IST-DRCU-ON-STEP2.tcl	—	—	—	
6	Check that THSK parameter is again refreshing every 4 seconds	—	—	—	
7	Check that TM2N parameter is again incrementing every 4 seconds	—	—	—	

Test Result (Pass/Fail):

Final Configuration:

- SPIRE DPU and DRCU are both on
- HK generation is on

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 6 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.3 Procedure SPIRE-IST-FUNC-SCU-02

Version: 1.0

Date: 12th June 2006

Purpose: SCU science contents check by the I-EGSE

Duration: 5 minutes

Preconditions:

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Procedure Steps:

Step	Description	Parameter	Expected Values Before/ After	Actual Values Before /After	Success/ Failure
1	Execute TCL script SPIRE-IST-FUNC-SCU-02.tcl	SCUFRAMECNT TM5N	31/62 1/3		
2	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 7 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.4 Procedure SPIRE-IST-FUNC-SCU-03

Version: 1.0

Date: 21st June 2006

Purpose: SCU DC thermometry check

Duration: 6 minutes

Preconditions: SPIRE FM is electrically integrated with the Herschel Satellite

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Execute TCL script SPIRE-IST-FUNC-SCU-03.tcl	—	—	—	
2	Wait for the parameter BBFULLTYPE to get set to SCU_DC_Therm	BBFULLTYPE	- /SCU_DC_Therm/ -		
3	A few seconds later record the value of parameter SCUTEMPSTAT	SCUTEMPSTAT	0/0xFFFF/0xFFFF		
4	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: SCU DC thermometry is switched on.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 8 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.5 Procedure SPIRE-IST-FUNC-SCU-06

Version: 1.0

Date: 12th June 2006

Purpose: SCU AC thermometry check

Duration: 2 minutes

Preconditions: SPIRE FM is electrically integrated with the Herschel Satellite

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Execute TCL script SPIRE-IST-FUNC-SCU-06.tcl	—	—	—	
2	Wait for the parameter BBFULLTYPE to get set to SCU_AC_Therm	BBFULLTYPE	- /SCU_AC_Therm /-		
3	A few seconds later record the value of parameter SUBKSTAT	SUBKSTAT	0/1/1		
4	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: SCU AC thermometry is switched on.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 9 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.6 Procedure SPIRE-IST-FUNC-SCU-07

Version: 1.0

Date: 11th June 2005

Purpose: SCU cooler heaters check

Duration: 3 minutes

Preconditions: SPIRE FM is electrically integrated with the Herschel Satellite

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Execute TCL script SPIRE-IST-FUNC-SCU-07.tcl	—	—	—	
2	Wait for the parameter BBFULLTYPE to get set to Cooler_Htr_Chk	BBFULLTYPE	Cooler_Htr_Chk		
3	A few seconds later record the value of parameter EVHSV – the Evaporator Heat Switch Voltage. <i>This voltage stays on for ~20 seconds.</i>	EVHSV - mV	0/~323/0		
4	A few seconds after the EVHSV parameter has been set back to 0, record the value of parameter SPHSV – the Sorption Pump Heat Switch Voltage. <i>This voltage stays on for ~20 seconds.</i>	SPHSV - mV	0/~323/0		
5	A few seconds after the SPHSV parameter has been set back to 0, record the value of parameter SPHTRV – the Sorption Pump Heater Voltage. <i>This voltage stays on for ~20 seconds.</i>	SPHTRV - V	0/~8.8/0		
6	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 10 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.7 Procedure SPIRE-IST-FUNC-SCU-04

Version: 1.0

Date: 12th June 2006

Purpose: SCU PCAL check

Duration: 2 minutes

Preconditions: SPIRE FM is electrically integrated with the Herschel Satellite

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter Name - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Execute TCL script SPIRE-IST-FUNC-SCU-04.tcl The expected values during the test should be monitored when parameter BBFULLTYPE in the FUNCTIONAL TEST PARAMETERS display is set to PCAL_Check This usually happens about 30 seconds from the start of test execution.	PCALCURR - mA PCALV – V BBFULLTYPE	0.0/0.1/0.0 0.0/0.026/0.0 PCAL_Check		
2	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 11 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.8 Procedure SPIRE-IST-FUNC-PCAL-01

Version: 1.0

Date: 12th June 2006

Purpose: SCU PCAL characterisation

Duration: 8 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter Name - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Execute TCL script SPIRE-IST-FUNC-PCAL-01.tcl	—	—	—	—
2	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 12 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.9 Procedure SPIRE-IST-FUNC-SCU-05

Version: 1.0

Date: 21st June 2006

Purpose: SCU Spectrometer SCAL4 and SCAL2 check

Duration: 4 minutes

Preconditions: SPIRE FM is electrically integrated with the Herschel Satellite

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Execute TCL script SPIRE-IST-FUNC-SCU-05.tcl	---	---	---	
2	Wait for the parameter BBFULLTYPE to get set to SCAL4_Check	BBFULLTYPE	SCAL4_Check		
3	A few seconds later record the value of parameters SCAL4CURR and SCAL4V <i>These parameters are set back to 0 after ~30 seconds</i>	SCAL4CURR – mA SCAL4V – V	0.0/0.10/0.0 0.0/0.05/0.0		
4	Wait for the parameter BBFULLTYPE to get set to SCAL2_Check	BBFULLTYPE	SCAL2_Check		
5	A few seconds later record the values of parameters SCAL2CURR and SCAL2V <i>These parameters are set back to 0 after ~30 seconds</i>	SCAL2CURR – mA SCAL2V – V	0.0/0.10/0.0 0.0/0.05/0.0		
6	Wait for the I-EGSE staff to confirm the success or failure of this test	---	---	---	---

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 13 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.10 Procedure SPIRE-IST-FUNC-SCAL-01

Version: 1.0

Date: 12th June 2006

Purpose: SCU SCAL4 and SCAL2 characterisation

Duration: 16 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter Name - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Execute TCL script SPIRE-IST-FUNC-SCAL-01.tcl	—	—	—	—
2	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 14 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.11 Procedure SPIRE-IST-FUNC-MCU-01

Version: 1.0

Date: 12th June 2006

Purpose: To boot up the MCU

Duration: 3 minutes

Preconditions:

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is DRCU_ON	MODE	DRCU_ON		
2	Execute TCL script SPIRE-IST-FUNC-MCU-01.tcl	—	—	—	
3	Check that the mode parameter is REDY	MODE	DRCU_ON/- /REDY		
4	Check that the MCU is booted up successfully	MCUBITSTAT	0/1/1		
5	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration:

- MCU is switched on and booted up
- SPIRE is in REDY mode

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 15 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.12 Procedure: SPIRE-IST-FUNC-MCU-03

Version: 1.0

Date: 12th June 2006

Purpose: MCU science data contents check

Duration: 5 minutes

Preconditions:

- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY	REDY	
2	Execute TCL script SPIRE-IST-FUNC-MCU-03.tcl	—	—	—	—
3	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 16 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.13 Procedure SPIRE-IST-FUNC-BSM-01

Version: 1.0

Date: 12th June 2006

Purpose: BSM switch on check

Duration: 3 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY		
2	Execute TCL script SPIRE-IST-FUNC-BSM-01.tcl	—	—	—	
3	Check that the Chop and Jiggle sensors have switched on	CHOPSENSPWR JIGGSENSPWR	0/1/1 0/1/1		
4	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: BSM is switched on.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 17 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.14 Procedure SPIRE-IST-FUNC-BSM-03

Version: 1.0

Date: 12th June 2006

Purpose: BSM open loop dynamics check

Duration: 6 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY		
2	Execute TCL script SPIRE-IST-FUNC-BSM-03.tcl	---	---	---	
3	Wait for the I-EGSE staff to confirm the success or failure of this test	---	---	---	---

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 18 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.15 Procedure SPIRE-IST-FUNC-BSM-05a

Version: 1.0

Date: 12th June 2006

Purpose: BSM open loop chop test

Duration: 10 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY		
2	Execute TCL script SPIRE-IST-FUNC-BSM-05a.tcl	—	—	—	
3	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 19 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.16 Procedure SPIRE-IST-FUNC-BSM-05b

Version: 1.0

Date: 12th June 2006

Purpose: BSM closed loop chop test

Duration: 10 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter – Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY		
2	Execute SPIRE-IST-BSM-INIT.tcl	—	—		
3	Execute TCL script SPIRE-IST-FUNC-BSM-03.tcl	—	—		
4	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 20 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.17 Procedure SPIRE-IST-BSM-OFF

Version: 1.0

Date: 12th June 2006

Purpose: Switch off the BSM

Duration: 2 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter – Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY		
2	Execute SPIRE-IST-BSM-OFF.tcl	—	—	—	
3	Check that the power to the BSM sensors is switched off	CHOPSENSPWR JIGGSENSPWR	1/-/0 1/-/0		
4	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: BSM is switched off.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 21 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.18 Procedure SPIRE-IST-FUNC-SMEC-02a

Version: 1.0

Date: 12th June 2006

Purpose: Open the SMECm launch latch

Duration: 5 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter – Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY		
2	Execute TCL script SPIRE-IST-FUNC-SMEC-02a.tcl	—	—	—	
3	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: SMECm is unlatched.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 22 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.19 Procedure SPIRE-IST-FUNC-SMEC-01

Version: 1.0

Date: 12th June 2006

Purpose: SMECm switch on check

Duration: 5 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter – Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY		
2	Execute TCL script SPIRE-IST-FUNC-SMEC-01.tcl	—	—	—	
3	Check that power to the SMEC LED and LVDT sensor is on	SMECENCPWR SMECLVDTPWR	0/-/4(TBD) 0/1/1		
4	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: SMECm is switched on.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 23 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.20 Procedure SPIRE-IST-FUNC-SMEC-03

Version: 1.0

Date: 12th June 2006

Purpose: SMEC LED Optical Encoder LED Check

Duration: 5 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter – Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY		
2	Execute TCL script SPIRE-IST-FUNC-SMEC-03.tcl	—	—	—	
3	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 24 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.21 Procedure SPIRE-IST-FUNC-SMEC-04a

Version: 1.0

Date: 12th June 2006

Purpose: SMECm open loop position test

Duration: 10 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter – Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY		
2	Execute TCL script SPIRE-IST-FUNC-SMEC-04a.tcl	—	—	—	
3	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 25 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.22 Procedure SPIRE-IST-FUNC-SMEC-09

Version: 1.0

Date: 12th June 2006

Purpose: SMECm open loop scan test

Duration: 10 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter – Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY		
2	Check the SMECm is in open loop	SMECLOOPMODE	6/6/6		
3	Execute TCL script SPIRE-IST-FUNC-SMEC-09.tcl	—	—	—	
4	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 26 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.23 Procedure SPIRE-IST-FUNC-SMEC-04b

Version: 1.0

Date: 12th June 2006

Purpose: SMECm closed loop position test

Duration: 10 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter – Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY		
2	Execute TCL script SPIRE-IST-FUNC-SMEC-INIT.tcl	SMECLOOPMODE	6/1/1		
3	Execute TCL script SPIRE-IST-FUNC-SMEC-04B.tcl	—	—	—	
4	Check that SMECm is still in closed loop	SMECLOOPMODE	1		
5	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: The SMECm is in closed loop.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 27 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.24 Procedure SPIRE-IST-FUNC-SMEC-07

Version: 1.0

Date: 12th June 2006

Purpose: SMECm closed loop scan test

Duration: 10 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter – Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY		
2	Execute TCL script SPIRE-IST-FUNC-SMEC-INIT.tcl	SMECLOOPMODE	-/1/1		
3	Execute TCL script SPIRE-IST-FUNC-SMEC-07.tcl	—	—	—	
4	Check that SMECm is still in closed loop	SMECLOOPMODE	1		
5	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 28 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.25 Procedure SPIRE-IST-FUNC-SMEC-06

Version: 1.0

Date: 12th June 2006

Purpose: SMECm saw-tooth test

Duration: 10 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter – Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY		
2	Execute TCL script SPIRE-IST-FUNC-SMEC-INIT.tcl	SMECLOOPMODE	-/1/1		
3	Execute TCL script SPIRE-IST-FUNC-SMEC-06.tcl	—	—	—	
4	Check that SMECm is still in closed loop	SMECLOOPMODE	1		
5	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 29 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.26 Procedure SPIRE-IST-SMEC-OFF

Version: 1.0

Date: 12th June 2006

Purpose: Switch off the SMEC

Duration: 2 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter – Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY		
2	Execute SPIRE-IST-SMEC-OFF.tcl	—	—	—	
3	Check that the power to the SMEC sensors is switched off	SMECENCPWR SMECLVDTPWR	4(TBD)/-/0 1/-/0		
4	Wait for the I-EGSE staff to confirm the success or failure of this test	—			

Test Result (Pass/Fail):

Final Configuration: SMECm is switched off.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 30 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.27 Procedure SPIRE-IST-FUNC-SMEC-02b

Version: 1.0

Date: 12th June 2006

Purpose: Close the SMECm launch latch

Duration: 2 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter – Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY		
2	Execute TCL script SPIRE-IST-FUNC-SMEC-02b.tcl	—	—	—	
3	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: SMECm is latched.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 31 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.28 Procedure SPIRE-IST-FUNC-DCU-01

Version: 1.0

Date: 12th June 2006

Purpose: DCU science packet generation check for all Photometer and Spectrometer packet types (PF, PSW, PMW, PLW, SF, SSW and SLW)

Duration: 5 minutes

Preconditions:

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter	Expected Values Before/After	Actual Values Before/After	Success/Failure
1	Execute TCL script SPIRE-IST-FUNC-DCU-01.tcl	DCUFRAMECNT	0/700		

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 32 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.29 Procedure SPIRE-IST-FUNC-DCU-11-P

Version: 1.0

Date: 12th June 2006

Purpose: Photometer detectors switch on

Duration: 10 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter – Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that Photometer LIAs are switched on	PLIABITSTAT	1		
2	Execute TCL script SPIRE-IST-FUNC-DCU-13-P.tcl	—	—	—	
3	Check that the Photometer detectors are switched on	PSWJFETSTAT PMLWJFETSTAT PSWJFET1V PSWJFET2V PSWJFET3V PSWJFET4V PSWJFET5V PSWJFET6V PMWJFET1V PMWJFET2V PMWJFET3V PMWJFET4V PLWJFET1V PLWJFET2V TCJFETV PHOTHTRV	0/0x3F/0x3F 0/-/0x7F -1.49V -1.49V -1.49V -1.49V -1.49V -1.49V -1.49V -1.49V -1.49V -1.49V -1.49V -1.49V -1.49V -1.49V -1.49V 0.0V		
4	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Photometer detectors are switched on.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 33 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.30 Procedure SPIRE-IST-FUNC-DCU-13-P

Version: 1.0

Date: 12th June 2006

Purpose: Perform a Photometer Load Curve

Duration: 20 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- Procedure SPIRE-IST-FUNC-DCU-11-P has been executed

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- The Photometer detectors and LIAs are on
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that Photometer LIAs are switched on	PLIABITSTAT	1		
2	Execute TCL script SPIRE-IST-FUNC-DCU-13-P.tcl	—	—	—	
3	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 34 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.31 Procedure SPIRE-IST-PDET-OFF

Version: 1.0

Date: 12th June 2006

Purpose: Switch off photometer detectors

Duration: 2 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- SPIRE-IST-FUNC-DCU-11-P has been executed

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- Photometer detectors and LIAs are switched on
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Execute TCL script SPIRE-IST-PDET-OFF.tcl	---	---	---	
2	Check that the Photometer detectors are switched off	PSWJFETSTAT PMLWJFETSTAT PSWJFET1V PSWJFET2V PSWJFET3V PSWJFET4V PSWJFET5V PSWJFET6V PMWJFET1V PMWJFET2V PMWJFET3V PMWJFET4V PLWJFET1V PLWJFET2V TCJFETV PHOTHTRV	0x3F/-/0 0x7F/-/0 0.0V 0.0V 0.0V 0.0V 0.0V 0.0V 0.0V 0.0V 0.0V 0.0V 0.0V 0.0V 0.0V 0.0V 0.		
3	Check that the Photometer LIAs are switched off	PLIABITSTAT	1/-/0		
4	Wait for the I-EGSE staff to confirm the success or failure of this test	---	---	---	---
Test Result (Pass/Fail):					

Final Configuration: Photometer detectors are switched off.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 35 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.32 Procedure SPIRE-IST-FUNC-DCU-11-S

Version: 1.0

Date: 12th June 2006

Purpose: Spectrometer detectors switch on

Duration: 10 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter – Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that Spectrometer LIAs are switched on	SLIABITSTAT	1		
2	Execute TCL script SPIRE-IST-FUNC-DCU-11-S.tcl	—	—	—	
3	Check that the Spectrometer detectors are switched on	SPECJFETSTAT SSWJFET1V SSWJFET2V SLWJFET1V SPECHTRV	0/7/7 -1.49V -1.49V -1.49V 0.0V		
4	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Spectrometer detectors switched on.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 36 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.33 Procedure SPIRE-IST-FUNC-DCU-13-S

Version: 1.0

Date: 15th June 2006

Purpose: Perform a Spectrometer Load Curve

Duration: 20 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- Procedure SPIRE-IST-FUNC-DCU-11-S has been executed

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- Spectrometer detectors and LIAs are switched on
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that Spectrometer LIAs are switched on	SLIABITSTAT	1		
2	Execute TCL script SPIRE-IST-FUNC-DCU-13-S.tcl	—	—	—	
3	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Unchanged

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 37 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.34 Procedure SPIRE-IST-SDET-OFF

Version: 1.0

Date: 12th June 2006

Purpose: Switch off Spectrometer detectors

Duration: 2 minutes

Preconditions:

- SPIRE FM is electrically integrated with the Herschel Satellite
- Procedure SPIRE-IST-FUNC-DCU-11-S has been executed

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- Spectrometer detectors and LIAs are switched on
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Execute TCL script SPIRE-IST-SDET-OFF.tcl	—	—	—	
2	Check that the Spectrometer detectors are switched off	SPECJFETSTAT SSWJFET1V SSWJFET2V SLWJFET1V SPECHTRV	7/-/0 0.0V 0.0V 0.0V 0.0V		
3	Check that the Spectrometer LIAs are switched off	SLIABITSTAT	1/-/0		
4	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: Spectrometer detectors are switched off.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 38 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.35 Procedure SPIRE-IST-MCU-OFF

Version: 1.0

Date: 12th June 2006

Purpose: Switch off the MCU – if necessary

Duration: 2 minutes

Preconditions:

- SPIRE is in REDY mode

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- SPIRE FUNCTIONAL PARAMETERS display is selected on the CCS

Step	Description	Parameter – Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Check that the mode parameter is REDY	MODE	REDY	REDY	
2	Execute SPIRE-IST-MCU-OFF.tcl	—	—	—	
3	Check that the MCU is switched off	MCUBITSTAT	1/-/0		
4	Wait for the I-EGSE staff to confirm the success or failure of this test	—			

Test Result (Pass/Fail):

Final Configuration: MCU switched off.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 39 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.36 Procedure SPIRE-IST-SCU-OFF

Version: 1.1

Date: 12th June 2006

Purpose: Switch off SCU DC and AC thermometry – if necessary

Duration: 2 minutes

Preconditions:

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Execute TCL script SPIRE-IST-SCU-OFF.tcl	—	—	—	
2	A few seconds later record the value of parameter SCUTEMPSTAT	SCUTEMPSTAT	FFFF/-/0		
3	A few seconds later record the value of parameter SUBKSTAT	SUBKSTAT	1/-/0		
4	Check that SPIRE is in DRCU_ON mode	MODE	REDY/-/DRCU_ON		
5	Wait for the I-EGSE staff to confirm the success or failure of this test	—	—	—	—

Test Result (Pass/Fail):

Final Configuration: SPIRE in DRCU_ON mode.

SPIRE Procedure

Ref: SPIRE-RAL-PRC-002398

Issue: 2.0

Date: 12/06/2006

Page: 40 of 41

SPIRE IST Cold Functional Test Procedures
S.D.Sidher & A.A.Aramburu

2.37 Procedure SPIRE-IST-DRCU-OFF

Version: 1.0

Date: 12th June 2006

Purpose: Switch off the DRCU – if necessary

Duration: 2 minutes

Preconditions:

- Procedure SPIRE-IST-SCU-OFF has been successfully executed
- SPIRE is electrically integrated with the Herschel EQM.

Initial Configuration:

- SPIRE DPU is on and generating HK
- DRCU is switched ON
- FUNCTIONAL TEST PARAMETERS display is selected on the CCS

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	Execute TCL script SPIRE-IST-DRCU-OFF.tcl	—	—	—	
2	Check that THSK parameter is not refreshing anymore	—	—	—	
3	Check that TM2N parameter is not incrementing anymore	—	—	—	
4	When instructed by the I-EGSE staff Power off the SPIRE DRCU using CCS procedure XXXXXX	—	—	—	

Test Result (Pass/Fail):

Final Configuration:

- DRCU is switched off
- SPIRE DPU is on but not generating HK



Report

SPIRE HCSS v0.3.3 Acceptance Test Report
Sunil Sidher, Steve Guest & Asier Aramburu

Ref:	SPIRE-RAL-REP-002587
Issue:	1.0
Date:	2 nd March 2006
Page:	41 of 41

2.38 Procedure SPIRE-IST-DPU-OFF

Version: 1.0

Date: 12th June 2006

Purpose: Switch off the DPU – if necessary

Duration: 2 minutes

Preconditions: SPIRE-IST-DRCU-OFF has been successfully executed.

Initial Configuration:

- SPIRE DPU is on *but not* generating any HK
- DRCU is switched OFF

Step	Description	Parameter - Unit	Expected Values Before/ During/ After	Actual Values Before/ During/ After	Success/ Failure
1	When instructed by the I-EGSE staff Power off the SPIRE DRCU using the CCS procedure XXXXXX	—	—	—	
Test Result (Pass/Fail):					

Final Configuration: SPIRE DPU is switched off and the SPIRE instrument is OFF.