

SPIRE Science Verification Review - 2

RAL

September 26, 2006

Optical Performances

Document Number: SPIRE-RAL-REP-002572 2.0

Marc Ferlet (RAL)

Contents

1.	Introduction and scope	2
2.	List of requirements that the test programme was designed to evaluate	2
3.	Test results and conclusions	4
3.1	List of tests carried out and tests still to be done	4
3.2	Subsystem requirements tested at instrument level and their verification status	4
3.3	Instrument-level requirements and their verification status	6
4.	Open issues and anomalies	7
5.	Recommendations for further data analysis and test	8
5.1	Recommended data analysis	8
5.2	Recommended tests	9
6.	References	11
6.1	Applicable documents	11
6.2	Reference documents	11

1. Introduction and scope

This document summarises the present status of the verification of the SPIRE instrument optical performances. List of the relevant requirements is given in section 2. Section 3 gives the principal results obtained so far with respect to individual optics subsystem level and optics-related instrument level requirements. A few anomalies found are reported in section 4. Some recommendations for priority data analysis and future tests are discussed in section 5. Finally, references for the background technical documentation are to be found in section 6.

The document template and content are identical to the first SVR document with targeted updates highlighted in blue text like this one.

2. List of requirements that the test programme was designed to evaluate

The optics-related requirements to be evaluated are listed in the table below; along side the relevant SPIRE model which has allowed/will allow their partial or complete verification. The requirement identifiers are taken from AD1.

Requirement Name	Description	Verification Method	Model	Test ID	Upper Links
IRD-OPTP-R00	Compatibility with Herschel telescope	Design analysis Instrument level alignment verification Instrument level performance tests	AM CQM	ILT_ALIGN ILT_PERF	IID-A-SECT4.3.1
IRD-OPTP-R01	Nominal final focal ratio	Design analysis	N/A	N/A	
IRD-OPTP-R02	Variation in focal ratio	Design analysis Instrument level alignment verification Instrument level performance tests	AM CQM PFMI PFMII	ILT_ALIGN ILT_PERF	IRD-PHOT-R10
IRD-OPTP-R03	Distortion	Design analysis Instrument level alignment verification Instrument level performance tests	AM CQM PFMI PFMII	ILT_ALIGN ILT_PERF	IRD-PHOT-R10
IRD-OPTP-R04	Anamorphism	Design analysis Instrument level alignment verification Instrument level performance tests	AM CQM PFMI PFMII	ILT_ALIGN ILT_PERF	IRD-PHOT-R03
IRD-OPTP-R06	Image quality	Design analysis Instrument level alignment verification Instrument level performance tests	AM CQM PFMI PFMII	ILT_ALIGN ILT_PERF	IRD-PHOT-R03 IRD-PHOT-R04 IRD-PHOT-R05
IRD-OPTP-R08	In-band straylight	Design analysis Instrument level performance tests	CQM PFMI PFMII	ILT_PERF	IRD-PHOT-R04
IRD-OPTS-R01	Nominal final focal ratio	Design analysis	N/A	N/A	
IRD-OPTS-R02	Variation in focal ratio	Design analysis Instrument level alignment verification Instrument level performance tests	AM CQM PFMI PFMII	ILT_ALIGN ILT_PERF	
IRD-OPTS-R03	Distortion	Design analysis Instrument level alignment verification Instrument level performance tests	AM CQM PFMI PFMII	ILT_ALIGN ILT_PERF	
IRD-OPTS-R04	Anamorphism	Design analysis Instrument level alignment verification Instrument level performance tests	AM CQM PFMI PFMII	ILT_ALIGN ILT_PERF	IRD-SPEC-R05
IRD-OPTS-R06	Image quality	Design analysis Instrument level	AM	ILT_ALIGN	IRD-SPEC-R05

Requirement Name	Description	Verification Method	Model	Test ID	Upper Links
		alignment verification Instrument level performance tests	CQM PFMI PFMII	ILT_PERF	IRD-SPEC-R06 IRD-SPEC-R07
IRD-DETS-R07	Detector angular response	Design analysis Subsystem acceptance data package	N/A	N/A	IRD-SPEC-R05
IRD-DETP-R07	Detector angular response	Design analysis Subsystem acceptance data package	N/A	N/A	IRD-PHOT-R06
IRD-STRC-R01	Alignment of the instrument w.r.t. the FIRST optical axis	Design analysis Instrument alignment verification	AM CQM	ILT_ALIGN	IID-A-SECT5.3.2.1
IRD-STRC-R04	Optics and associated sub-system alignment	Design analysis Instrument alignment verification	AM CQM	ILT_ALIGN	
IRD-STRC-R08	Attenuation of radiation from cryostat environment	Design analysis Instrument level performance tests	CQM	ILT_PERF	IRD-PHOT-R04 IRD-PHOT-R05 IRD-SPEC-R17
IRD-STRP-R02	Optics and filters alignment	Design analysis Instrument alignment verification	AM CQM	ILT_ALIGN	
IRD-STRP-R03	Array module alignment	Design analysis Instrument alignment verification	AM CQM	ILT_ALIGN	IRD-PHOT-R16
IRD-STRP-R06	Attenuation of radiation from common structure environment	Design analysis Instrument level performance tests	CQM	ILT_PERF	IRD-PHOT-R04 IRD-PHOT-R05
IRD-STRS-R02	Optics alignment requirements	Design analysis Instrument alignment verification	AM CQM	ILT_ALIGN	
IRD-STRS-R03	Array module alignment	Design analysis Instrument alignment verification	AM CQM	ILT_ALIGN	
IRD-STRS-R06	Attenuation of radiation from 4-K environment	Design analysis Instrument level performance tests	CQM	ILT_PERF	IRD-SPEC-R17
IRD-PHOT-R02	FoV	Design analysis Instrument level alignment verification Instrument level performance tests	CQM PFM I PFM II	ILT_ALIGN ILT_PERF	SRD R7
IRD-PHOT-R03	Beam FWHM	Design analysis Instrument level Instrument level performance tests	CQM PFM II	ILT_PERF	SRD R1
IRD-PHOT-R10	Field distortion	Design analysis Instrument level alignment verification Instrument level performance tests	AM CQM PFM I PFM II	ILT_ALIGN ILT_PERF	SRD R6
IRD-PHOT-R16	Co-alignment	Design analysis Instrument level alignment verification Instrument level performance tests	AM PFM I PFM II	ILT_ALIGN ILT_PERF	SRD R15
IRD-SPEC-R04	FoV	Design analysis Instrument level performance tests	PFM I PFM II	ILT_PERF	SRD R16
IRD-SPEC-R05	Beam FWHM	Design analysis Instrument level performance tests	PFM I PFM II	ILT_PERF	SRD R16

3. Test results and conclusions

3.1 List of tests carried out and tests still to be done

Tests relevant to optical alignment (ILT_ALIGN in table above), mostly for AM and CQM models are discussed in RD1 and RD2.

Test relevant to in-band optical performances (ILT_PERF in table above) test on CQM and PFM1 are discussed in RD4 and RD5 with support from the specific characteristics of the optical set-up of the SPIRE test facility, described in RD3.

More details on the SPIRE test campaign details (plan, test data, intermediate reports, ...) on the can be found at the following link: http://scott1.bnsc.rl.ac.uk:8080/hcss/test_area/index.htm .

3.2 Subsystem requirements tested at instrument level and their verification status

The table below summarised the optics subsystem requirements and their present status.
Updates from PFM3 have been put again in blue where relevant.

Requirement Name	Description	Verification status
IRD-OPTP-R00	Compatibility with Herschel telescope	Nominally by design HSO telescope taken HSO axis and pupil simulated during AM warm and cold alignment activities as reference for pupil alignment (internal/external).
IRD-OPTP-R01	Nominal final focal ratio	Taken as F/5 by design. Not derived (yet) from measurement. ⇒ Update from PFM3: preliminary derived value for Phot arrays is ~4.9+/-0.25. Refined value and reduction of uncertainty from future cross-check with other datasets.
IRD-OPTP-R02	Variation in focal ratio	Ok by design. To be derived from future Phot full FoV mapping and PSF measurements ⇒ Update from PFM3: see above. Presently compliant and to be refined.
IRD-OPTP-R03	Distortion	Ok by design. To be derived from future Phot full FoV mapping measurement ⇒ Update from PFM3: preliminary derived value for Phot arrays is <8+2/-5%. Refined value and reduction of uncertainty from future cross-check with other datasets.
IRD-OPTP-R04	Anamorphism	Ratio of axes length in assumed generally elliptical beam pattern found <~10% for PLW from CQM (at FoV centre) so compliant to the 6/5 max ratio ; to be confirmed by PFM2 PLW measured beam data. To be analysed for PSW. PMW affected by the anomaly during PFM2 so possibly not relevant. ⇒ Update from PFM3 (& PFM2): data analysis on PLW confirmed the compliance at <10% level. Value <15% for PSW and ~5% for PMW. Obtained by max ellipticity of pixel spatial response for in-band point source on a sample set of pixels.
IRD-OPTP-R06	Image quality	Strehl ratio >99% at 250um over full Phot FoV; derived from measurement during AM warm alignment. Not derived (yet) from in-band measurement (through-focus). ⇒ Update from PFM3: derivation of in-band Strehl ratios from spectral measurements of spatial pixel response and modelling of low Fresnel number optical effect. Values ~98+/-2% for PSW, 96+/-2% for PMW and 94+/-3% for PLW have been obtained so compliant with the >0.9 spec. Apparent anomaly of the decrease with wavelength explained by optical diffractive effect involved and measurement performed at best focus for individual wavelength. Refined values to be obtained by cross-checking with other test set (see recommended tests for PFM4) and alternative analysis method.

Requirement Name	Description	Verification status
IRD-OPTP-R08	In-band straylight	To be discussed in the spectral & throughput performances document. ⇒ Analysis of waveguide cut-off and from filter/dichroics explained the measured spatial response beyond the nominal band edges of PMW and PSW in PFM2. Qualitative appearance of ghost images in PFM3 data when external point source used in the measurement but can not separated from the ones induced by extra test cryostat filter, so final status of Phot in-field in-band straylight to be obtained by re-analysis. Out-of-field in-band straylight can only be commented in the throughput and photometry document as not appearing in optical data set (no source modulation).
IRD-OPTS-R01	Nominal final focal ratio	Taken initially as ~F/5 but final design gives ~4.85+/-0.1 for SLW and 4.35+/-0.1 for SSW (uncertainty is variation across FoV). From PFM1: measured to be between 4.5 and 5 for SLW; measured to be between 4 and 5 for SSW. ⇒ No update from PFM3 due to reduced/partial coverage of SSW FoV only (half not accessible as not working); pixel characterisation of SLW only not FoV.
IRD-OPTS-R02	Variation in focal ratio	From above: Ok by design and ok as derived from PFM measurements. ⇒ See IRD-OPTS-R02 above
IRD-OPTS-R03	Distortion	Ok by design; estimated to <~5% from field mapping. ⇒ See IRD-OPTS-R02 above
IRD-OPTS-R04	Anamorphism	From PFM1: found <~10% for the all pixels tested in SSW and SLW. Some uncertainty remains due to test source fluctuations and shape at different wavelengths. ⇒ Update from PFM3: value from PFM1 confirmed. Slight dependence over the FoV noticed for SLW (lower anamorphism closer to FoV centre); not clearly the case for SSW.
IRD-OPTS-R06	Image quality	Strehl ratio >95% at 250um at FoV centre; derived from measurement during AM warm alignment. Improved to >97% after astigmatism correction in Spec optical train for PFM. Not derived (yet) from in-band measurement (through-focus). ⇒ No update from PFM3 as no through-focus data set for Spectro but same phenomenon as in IRD-OPTP-R06 expected (but more difficult to extract separately from the pixel response more dominated by detector modal behaviour in the Spectro).
IRD-DETS-R07	Detector angular response	By design in Spec BDAs EIDP Modal content to be further investigated (see section 5.2) at different wavelengths and correlated with spectral bandpass measurement ⇒ Update from PFM3: quasi single-mode behaviour retrieved and confirmed for SSW; quasi single mode for SLW for $\lambda > 580\mu\text{m}$ and multi-moded between 300um and ~450um; obtained from experimental derivation of in-band etendue (radiometric; in agreement with geometric so optics/detector interface) at set of measured in-band wavelength.
IRD-DETP-R07	Detector angular response	By design in Phot BDAs EIDP Modal content to be further investigated (see section 5.2) at different wavelengths and correlated with spectral bandpass measurement ⇒ Update from PFM3: near single-mode behaviour retrieved and confirmed from experimental derivation of in-band etendue at set of measured in-band wavelength.
IRD-STRC-R01	Alignment of the instrument w.r.t. the FIRST optical axis	By FPU structure design & characteristics
IRD-STRC-R04	Optics and associated sub-system alignment	Test on AM warm and cold for pupil alignment (after vibrations) + warm CQM => ok within spec. Test warm on PFM (before vibrations) -> ok well within spec. No filters, BDAs, or BSM (OGSE flat mirror CM4 used during AM & CQM model test then used as reference for mounting flight BSM).
IRD-STRC-R08	Attenuation of radiation from	See FPU structure design & justification

SPIRE	SPIRE Science Verification Review - 2 Optical Performances	Ref: SPIRE-RAL-REP-002572 Issue: 2.0 Date: 15/09/06 Page: 6 of 12
--------------	---	--

Requirement Name	Description	Verification status
	cryostat environment	
IRD-STRP-R02	Optics and filters alignment	Test on AM warm and cold for pupil alignment (after vibrations) + warm CQM => ok within spec. Test warm on PFM (before vibrations) -> ok well within spec.
IRD-STRP-R03	Array module alignment	No filters, BDAs, or BSM (OGSE flat mirror CM4 used during AM & CQM model test then used as reference for mounting flight BSM). BDA alignment covered by BDA metrology fed into design of interface plates at detector box/BDA interfaces.
IRD-STRP-R06	Attenuation of radiation from common structure environment	See FPU structure design & justification
IRD-STRS-R02	Optics alignment requirements	Test on AM warm and cold for pupil alignment (after vibrations) => ok within spec. Test warm on PFM (before vibrations) => ok well within spec
IRD-STRS-R03	Array module alignment	No filters, BDAs, or BSM (OGSE flat mirror CM4 used during AM & CQM model test then used as reference for mounting flight BSM). BDA alignment covered by BDA metrology fed into design of interface plates at detector box/BDA interfaces.
IRD-STRS-R06	Attenuation of radiation from 4-K environment	See FPU structure design & justification

3.3 Instrument-level requirements and their verification status.

The table below summarised the higher level relevant requirements and their present status. Updates from PFM3 have been put again in blue where relevant.

Requirement Name	Description	Verification Status
IRD-PHOT-R02	FoV	4x8arcmin ² nominal from design. Used as-is during AM activities but not quantitatively assessed. Not assessed in-band (yet). ⇒ Update from PFM3: FoV found to be ~4x8arcmin ² for all Phot bands for spatial mapping of Fov via peak-ups
IRD-PHOT-R03	Beam FWHM	PSW: TBC (measured during PFM2 but final value not yet derived) ⇒ Update from PFM3 (and PFM2): 19.2+/-2.0arcsec when illumination with broadband thermal point source but extra refined value and relative uncertainty reduction when all PSW dataset taken into account in analysis ⇒ 18+/-1.4arcsec at 250um from the spectral average of response to wavelength-varying in-band source PMW: TBC but potential discrepancy due to PMW anomaly (see section 4) ⇒ Update from PFM3: 22.3+/-2.1arcsec when illumination with broadband thermal point source but extra refined value and relative uncertainty reduction when all PMW dataset taken into account in analysis ⇒ 23.5+/-1.6arcsec at 350um from the spectral average of response to wavelength-varying in-band source PLW: 30+/-4.8arcsec at 432um (measured during CQM2 tests), equivalent to 34.7+/-5.5 at 500um. Uncertainty to be reduced by analysis of PFM2 data. A priori, compliant wrt spec: 36arcsec at 500um ⇒ Update: PFM2 -> 35.0+/-2.0arcsec // PFM3 -> 36.5+0.5/-1.3arcsec when illumination with broadband thermal point source ⇒ 32.2+/-1.8arcsec at 500um from the spectral average of response to wavelength-varying in-band source

SPIRE	SPIRE Science Verification Review - 2 Optical Performances	Ref: SPIRE-RAL-REP-002572 Issue: 2.0 Date: 15/09/06 Page: 7 of 12
--------------	---	--

Requirement Name	Description	Verification Status
IRD-PHOT-R10	Field distortion	To be derived from future Phot full-Fov measurement. ⇒ Update from PFM3: found <10% (compliant) for all Phot bands for spatial mapping of Fov via peak-ups
IRD-PHOT-R16	Co-alignment	Not directly measured during AM activities as BDAs and filter not present. Used as design guideline in BDA/Phot detector box alignment hardware design (see SPIRE-RAL-NOT-002344 issue 1.0 (15/03/05)). Good general alignment (quantitatively TBC) found between PLW and PSW. PMW appeared not perfectly co-aligned during PFM2, likely due to the anomaly (see section 4). ⇒ Update from PFM3: found <10% (compliant) for all Phot bands for spatial mapping of Fov via peak-ups
IRD-SPEC-R04	FoV	2.6arcmin diameter nominal from design. Not used or assessed during AM activities (on-axis use only) + not present in CQM. Spec field mapping during PFM indicate FoV is at least 2.6arcmin (although small continuous vignetting occur right from the FoV centre) for both SSW and SLW. ⇒ No special update as PFM3 coverage of Spectro Fov was only partial
IRD-SPEC-R05	Beam FWHM	SSW: 16+/-2arcsec at 250um (extrapolated from value derived from PFM1 measurement) So compliance with spec of 18arcsec at 250um. ⇒ Update from PFM3: 13.8+/-1.1arcsec when illumination with broadband thermal point source ⇒ From the spectral average of response to wavelength-varying in-band source: variation from 17.4arcsec at ~200um down to 15.6arcsec at ~240um and back up to 17.3arcsec at 300um SLW: measured during PFM1 >40arcsec below ~310um and <30arcsec above 500um So not compliant with the spec of 25arcsec at 350um but the spec does not take into account the multi-modal behaviour of the detector+coupling optics (broadband waveguide + feedhorn). ⇒ Update from PFM3: 24.2+/-1.3 arcsec when illumination with broadband thermal point source ⇒ From the spectral average of response to wavelength-varying in-band source: variation from nearly 40arcsec at 300um down to ~31arcsec near 450/500um and back up to 35-40arcsec above 600um + high fluctuations of the value in the 300-450um region due to dominant multi-moded behaviour of the detector (feedhorn+waveguide+cavity)

NB: any value in arcsec on arcmin in the above table on-sky is extrapolated from experimentally derived data by assuming nominal baseline HSO telescope characteristics. Also the values mentioned at specific wavelength or for special source type have been measured at best respective focus (i.e. where best coupling via maximal signal from detector after centration of the point source on the given pixel).

4. Open issues and anomalies

- **Spatial response of PMW array:**

During PFM2 test campaign, it was observed a different spatial response. It acted as PMW channel was affected by field dependent aberrations inducing change in shape of the PSF from pixel to pixel alongside an estimated longer in-band relative depth-of-focus and relatively large (in fraction of pixel) misalignment wrt PSW and PLW arrays.

After inspection after the test campaign, a bent/curved dichroic (in reflection for PMW) was found. The design expects a flat reflecting surface while a possible powered element was present during the PFM2 tests in front of PMW BDA. This can qualitatively explain the effects seen.

More quantitative estimations (comparison measured data/optical simulations) will be performed to assure that no other PMW BDA specific effects are to be the cause. Meanwhile the dichroic in question has been replaced by a flat one at ambient.

⇒ **Completed (analysis and replacement of PDIC2) before start of PFM3 test campaign. PMW data from PFM3 test shows normal behaviour (spatial response) i.e. the cause of the PFM2 PMW anomaly has been understood and corrected.**

- **Global illumination of the Phot arrays:**

A non-uniform illumination pattern of the Phot arrays has been found. The issue under investigations in an attempt to discriminate the possible cause: individual detector response, test source field flatness, extra misalignment at operating conditions or internal vignetting of Phot optical path at FoV edge not previously found.

- ⇒ **Analysis via optical models of SCAL source illumination of the Spectro arrays has been performed and indicate non-uniform illumination pattern in qualitative agreement with test results. PCAL illumination of Phot array under investigation with optical models. Anomalous (non-uniform and shifted) array illumination by the CBB during PFM3 qualitatively traced to a tilt of the CBB assembly at FPU aperture interface: improved mounting of CBB assembly at FPU interface will be implemented for PFM4 test campaign.**

- **Spec Channel fringing:**

The presence of low and high frequency channel fringing in the Spec SSW and SLW interferogram signals was detected during PFM1 testing. It seems to appear when looking at cold, spatially and spectrally broad source (not found when bright monochromatic point sources are used).

In-field ghost images were expected and simulated in both SSW and SLW in particular due to the presence of the additional field lenses in front of both BDAs and could lead to some channel fringing. And modelling of some Spec characteristics have led to good matching of other features of Spec interferogram signal such as the baseline variation and decentring (from internal apodisation, vignetting and SMEC localised small decentre and tilt).

But here the exact source of the different channel fringing effects is not entirely clear: supplementary reflections detector/lens, OPD-dependent out-of-field stray (internal ?) path are among the alternative explanations.

Corrective action such as the development and use of an AR coating on the Spec field lenses will reduce the magnitude of the spatially overlapping ghost images and therefore reduce the effect of the channel fringing.

- ⇒ **New AR coating on Spectro lens for PFM3. Noticeable reduction in low frequency channel fringe indicating the effect of the new coating but appearance of new non-predicted ghost across Spectro FoV when external test source used: qualitatively linked to the presence of extra cryostat filter (for room background/load reduction).**
- ⇒ **High frequency channel fringing still present so possible causes listed above still valid. Important note: SMEC has been replaced by the flight item after PFM3 campaign; a possible link between remaining channel fringing and SMEC configuration could then be further identified in during PFM4.**

5. Recommendations for further data analysis and test

5.1 Recommended data analysis

- **PFM2 spatial data:**

Most point source beam scans and peak-ups, monochromatic and broadband, as well as pupil scan profiles at different wavelength are to be analysed in order to characterise the individual pixel spatial response (PSF shape for derivation of FWHM, anamorphism, etendue).

NB: The Phot arrays were only partially active so distortion and variation at the Phot FoV scale can not be retrieved from PFM2 test data set.

- ⇒ **Completed during PFM3 test campaign. Analysis procedure further applied to the measured PFM3 dataset.**

- **Through-focus beam scans:**

Further analysis of this data set could be used to retrieve any eventual residual in-band relative defocus between arrays (i.e. longitudinal co-alignment), within Phot and Spec respectively, as well as in-band Strehl ratio (instead of the wavelength scaling technique used at the moment from Warm Alignment image quality test in the VNIR), including filter and detectors, from the axial variations of measured focal plane irradiance.

NB: Optical phase retrieval (including the effect of the coupling detector-feeding optics) could be performed but not expected to give high due to large depth-of-focus and natural long-wavelength blurring

- ⇒ **Only one Phot focus test performed during PFM3, none on Spectro (see below recommended tests for PFM4). Strehl ratio derived by alternative method from spectral measurement of spatial pixel**

response at best focus and compared to axial irradiance model. Phase retrieval from dataset (based on focus diverse phase diversity or else) as an independent assessment to be developed.

- **Merging of pupil scans and beam scans data:**

When measured with same calibration source at same wavelength, subtracting the measured focal plane pixel beam pattern back projected on the external pupil from the pupil scan data in order to obtain an estimate of the telescope pupil illumination.

NB: This is limited for the moment to a reduced number of wavelengths (1 for Spec during PFM1, 2 for Phot during PFM2).

⇒ **Has been done for the one measured on Spectro (SSW). Alternative derivation of effective edge taper as function of wavelength has been derived from PFM3-measured spatial pixel response at different in-band wavelength for the specific case of PLW. Extension to PSW and PMW on-going.**

5.2 Recommended tests

First, a brief review of the tests recommended at SVR-1 is performed.

- **Systematic mapping of the Phot FoV:**

The availability of the complete live Phot arrays for PFM3 performance test campaign would allow the Phot FoV mapping (extent, distortion). This would be performed by systematic series (along chop and jiggle axes as well as FoV corners) of point source peak-ups, moved by/from external test set-up with recording of the source pointing.

NB: for Spec, this was done nearly completely during PFM1 performance test campaign.

⇒ **Completed during PFM3 test campaign.**

- **Extension of the spatial characterisation tests (beam scans & peak-up):**

Extension of the standard spatial beam scan test with monochromatic point source to more wavelengths in-band for Phot and Spec would allow a more spectrally continuous derivation of the main characteristics (PSF width and shape) in each band and compared to alternative derivation of the wavelength-dependent throughput (e.g. from spectral bandpass measurement).

This needs the use of several wavelengths per Spec or Phot bands when available by the calibration source (FIR laser). The use of low power but detectable spectrally tunable photomixing source, tested during PFM2, can complement the FIR laser at the longer wavelength regions (in PLW and SLW bands).

⇒ **This is mainly a modification of an existing test to allow higher information, wrt pixel spatial response, to be recorded. During PFM3, more type II beamscreens were performed and the use of type III (see Calibration and Test Plan) as sampling/duration trade-offs for peak-ups.**

- **Extension of the pupil scan:**

The pupil scan test is at the moment performed by scanning a simulated point source at the SPIRE external exit pupil (matching HSO telescope secondary) only in one direction (along Z_{esa}), horizontal in the test configuration. In order to extend the verification the illumination of this circular external pupil for both Spec and Phot, it is recommended to perform the test as well the pupil scan in the orthogonal direction (Y_{esa}). A simple scheme for the implementation of this manual test in the present test set-up is being investigated.

⇒ **This is mainly an extension of an existing test to allow higher information wrt effective external pupil illumination, to be recorded. Not performed during PFM3.**

- **Polarisation response:**

To avoid spurious effect from source already polarised such as the FIR laser and supplementary polarisation rotation induced by the test set-up optics, it is recommended to perform the polarisation test with the broadband calibration source (HBB), typically in point source mode towards co-aligned Spec or Phot pixels. Implementation and procedure have already been developed accordingly.

NB: This is not directly called by any optics or instrument level requirements but its interest lies in the potentiality that a change in response as function of the polarisation angle could inducing some modulation (eventually spatially-dependent) in in-band throughput.

⇒ **Although not directly linked with/called by IRD spec, this is a baselined test. Not performed during PFM3, so recommended for PFM4 (see below).**

A final set of recommended optical tests for the PFM4 test campaign is listed in the table below. Note that most of them are labelled medium or low priority as they are intended for further characterisation and not direct evaluation of compliance with IRD. This is has been submitted for inclusion in the PFM4 test plan.

Recommended tests	Purpose and Justification	Outline of procedure	Priority for PFM4 test campaign
Post-vibration verification test	<i>Quick estimation of eventual post-vibrations optical performances degradation by comparison with PFM1,2 and 3 datasets</i>	reduced sets of peak-up with HBB on a few pixels for each bands Phot and Spectro + max 1 beamscan type II per band with HBB or laser or photonic mixer	<u>Medium/high priority</u>
Pupil scan	<i>In-band verification of external/internal pupil alignment (optical dataset quality & validity) + complementary information wrt pupil illumination (edge taper) at given wavelength</i>	1 in SLW and 1 in PMW but some issue with lack of stable FIR laser line available in PMW band as only doable with strong source (= laser); with 2 nd dimension extension (slice along Y axis)	<u>Low priority</u>
Spatial focus test	<i>Through-focus characterisation; Inter-band relative defocus assessment; Database of spatial response wrt defocus</i>	with HBB on at least co-aligned Phot and co-aligned Spectro pixels with range beyond respective nominal depth-of-focus + same with 1 line (laser or photomixer) per band	<u>Medium priority</u> <u>Low priority</u>
Spectral focus test	<i>Characterisation of defocus effect on Spectrometer pixel response (coupling, spectral shift, ...)</i>	SMEC scans with HBB (and/or photomixer) at co-aligned central Spectro pixels with different defocus in TelSim. NB: needs effective SCAL nulling of room background	<u>Medium/Low priority</u>
Polarisation test	<i>Assessment of instrument polarisation response impact on photometry and eventually the second order vector effects</i>	with HBB on co-aligned near FoV centre pixels for Phot and Spectro	<u>Medium/Low priority</u>
Spectro 3D characterisation test	<i>Characterisation of spatial/spectral crossed effects by generation of a spatially-oriented 3D data cube</i>	beam scans on Spectro at different OPD/SMEC positions: could take between ~5h min to ~20h depending on OPD step size + open question if to be done with HBB on co-aligned or with less stable laser lines (trade-off is 1 with HBB and 1 with photomixer) NB: needs effective SCAL nulling of room background	<u>Medium/Low priority</u>
Aperture test	<i>Characterisation of in-band pixel coupling vs source size from quasi-point source to quasi-extended source</i>	TelSim pointed at central co-aligned Phot and then Spectro pixels with HBB source (with laser and photomixer possible, just need strong source): object aperture size is varied from quasi close to wide open	<u>Low priority</u>

6. References

6.1 Applicable documents

AD1 SPIRE-RAL-PRJ-00034 issue 1.3 (14/07/05)

6.2 Reference documents

RD1 SPIRE STM alignment test campaign: Phot Hartmann test, Spec Hartmann test, cold stop alignment verification (05/03) + SPIE vol 5487 pp448-459 (06/04)

RD2 SPIRE-RAL-NOT-001807 issue 2.0 (07/07/04)

RD3 SPIRE-RAL-NOT-002006 issue [3.0 \(25/08/06\)](#)

RD4 SPIRE-RAL-NOT-002211 draft 0.3 (23/02/05)

RD5 SPIRE-RAL-NOT-002460 issue [2.0 \(25/08/06\)](#)