



Ref.:HSO-CDF-EIDP-073

Issue: 1.0

Date:15th March 2005

Page 1 of 50

SPIRE - Photometer Calibration Source (PCal) - PFM

End Item Data Package (EIDP)

SPIRE - Photometer Calibration Source (PCal) - PFM

SPIRE Ref.: SPIRE-UCF-DOC-002797

Cardiff Ref.: HSO-CDF-EIDP-073 Issue 1.0

15th March 2005

Prepared by: Peter Hargrave

Cardiff SPIRE Technical Manager

Approved by: Ian Walker

Cardiff AIG Programme Manager

Distribution list

RAL	Eric Sawyer	Cardiff	Peter Ade
	Eric Clark		Matt Griffin
	Judy Long		Ian Walker
	Bruce Swinyard		
	Doug Griffin	ATC	Phil Parr-Burman

Astronomy Instrumentation Group,
Department of Physics & Astronomy,
University of Wales, Cardiff,
5 The Parade,
Cardiff
CF24 3YB
+44 (0)2920 876682

X:\Cardiff workpackages\Config	ured Documents\PCAL\PFM-build\Docu	ments\EIDP\PCAL PFM HSO-CDF-EIDP-073.	.doc

Change Record

Issue	Section	Date	Changes
1.0		15/03/05	First issue for DRB

Table of contents

SECTION 01 - SHIPPING DOCUMENTS	7
SECTION 02 - TRANSPORTATION, PACKING, HANDLING & INTEGRATION PROCEDURES	10
SECTION 03 - CERTIFICATE OF CONFORMANCE	11
SECTION 04 - QUALIFICATION STATUS LIST / COMPLIANCE MATRIX	12
SECTION 05 - TOP LEVEL DRAWINGS (INC. FAMILY TREE)	15
SECTION 06 - INTERFACE DRAWINGS	16
SECTION 07 - FUNCTIONAL, BLOCK & MECHANICAL DRAWINGS	18
SECTION 08 - ELECTRICAL CIRCUIT DIAGRAMS	24
SECTION 09 - AS BUILT CONFIGURATION ITEMS STATUS LIST	26
SECTION 10 – SERIALISED COMPONENTS LIST	27
SECTION 11 - LIST OF WAIVERS	27
SECTION 12 - COPIES OF WAIVERS	27
SECTION 13 - OPERATIONS MANUAL	28
SECTION 14 - HISTORICAL RECORD	
SPIRE - Photometer Calibration Source (PCal) - PFM - End Item Data Package (EIDP) Page 3	of 50

SECTION 15 - LOGBOOK / DIARY OF EVENTS	31
SECTION 16 - OPERATING TIME / CYCLE RECORD	31
SECTION 17 – CONNECTOR MATING RECORD	32
SECTION 18 – AGE SENSITIVE ITEMS RECORD	33
SECTION 19 – PRESSURE VESSEL HISTORY / TEST RECORD	33
SECTION 20 - CALIBRATION DATA RECORD	33
SECTION 21 - TEMPORARY INSTALLATION RECORD	35
SECTION 22 - OPEN WORK / DEFERRED WORK / OPEN TESTS	36
SECTION 23 - LIST OF NON-CONFORMANCE REPORTS	36
SECTION 24 - COPIES OF NON-CONFORMANCE REPORTS	36
SECTION 25 - TEST REPORTS	36
SECTION 26 – ASSEMBLY RECORD	36
SECTION 27 - REFERENCE LIST OF EIDP'S	37
SECTION 28 - MASS RECORDS	38
SECTION 29 - CLEANLINESS STATEMENT	39
SPIRE - Photometer Calibration Source (PCal) - PFM - End Item Data Package (EIDP) Page	e 4 of 50

SECTION 30 - OTHER USEFUL INFORMATION	40
SECTION 31 - DPL/DML	41
SECTION 32 – LIST OF APPENDICES/ATTACHMENTS	41
APPENDIX A	42
THERMOMETER CALIBRATION REPORTS	ERROR! BOOKMARK NOT DEFINED.
APPENDIX B	43
REPORT FROM CRYOGENIC VIBRATION TESTS	ERROR! BOOKMARK NOT DEFINED.
APPENDIX C	44
SCAL-B CALIBRATION REPORT	ERROR! BOOKMARK NOT DEFINED.
APPENDIX D	45
SCAL LIFETEST REPORT	45
APPENDIX E	46
SCAL LIFETEST PROCEDURE & RECORD	ERROR! BOOKMARK NOT DEFINED.
APPENDIX F	47
SCAL ENHANCED WARM-UP TEST REPORT	ERROR! BOOKMARK NOT DEFINED.
SPIRE - Photometer Calibration Source (PCal) - PFM - End Item Data Package (EIDP)	Page 5 of 50

PPENDIX G	
ERTIFICATES OF CONFORMANCE & CALIBRATION	ERROR! BOOKMARK NOT DEFINE
SPIRE - Photometer Calibration Source (PCal) - PFM - End Item I	Data Package (EIDP) Page 6 of 50

SECTION 01 - Shipping Documents

Minutes of SPIRE PCAL-PFM DRB – to be completed at DRB.....

Cardiff, 2005

Present:

DRB scope and objectives

- Review subsystem documentation, qualification and performance status of PFM PCAL unit
- Formal acceptance of the hardware

List of relevant documents and status

•

Review of comments received on EIDP

0

EIDP run-through

•

Conclusions and summary of outstanding work

DRB Check-list

If all listed points are satisfactory then the item may be accepted. If not, the item may be accepted with the caveat that the requirement in question will be met within an agreed time scale.

No.	Record applicable answers and provide explanations in comment column	Yes/No/NA	Comments
1	Are all documents, drawings, etc. complete, approved, and under configuration control?		
2	Do the released items above reflect all approved changes?		
3	Is the hardware identical to other hardware delivered. If not provide difference list?		
4	Does the hardware fulfil its functional requirements, specifications, RFWs, ICDs etc.?		
5	Does the As-built Configured Items List reflect hardware as delivered?		
6	Have all required environmental tests and analyses been completed?		
7	Are all the required test and qualification procedures and reports completed and available to review.		
8	Have all the Declared Lists, i.e. DML, DMPL, DPL and components (EEE Parts) list been released? (Need not necessarily be included in EIDP, but must be available)		
9	Have PADS been raised and approved where required?		
10	Are relevant Test Readiness Review (TRR) minutes and the Post Test		

	Reviews (PTR) complete and available?	
11	EIDP check: Are all agreed sections of the EIDP complete? Record any	
	anomalies.	

		Signature	Date
Ian Walker	Cardiff SPIRE PA manager		
Peter Hargrave	Cardiff SPIRE Technical Manager		
Matt Griffin	DRB chair		

SECTION 02 - Transportation, Packing, Handling & Integration Procedures

Handling

- Full ESD precautions to be observed when handling PCal.
- Inspection may be carried out in class-100 clean air cabinet.
- The rear cover must not be removed unless an authorised member of Cardiff personnel is present.

Storage

• PCal must be stored in the shipping container provided.

Installation

• Installation of PCal will be carried out by an ESA-approved soldering technician at RAL, according to the PCal ICD – HSO-CDF-ICD-013.

Cleaning & Bakeout

- Outer surface may be cleaned using a clean-room wipe impregnated with iso-propyl alcohol, with a piece of Kapton tape covering the aperture.
- Bakeouts should be kept to a minimum, and the temperature should NEVER exceed 90°C.

SECTION 03 - Certificate of Conformance

Spacecraft / Project:	Herschel	
Instrument:	SPIRE	
Model:	PFM	
Subsystem:	Photometer Calibration Source	
Serial No: PCAL-PFM-000		
As described in this End Item Data Package: HSO-CDI	F-EIDP-073	

Complies with the requirements set out in:
SPIRE Instrument Requirements Document - SPIRE-RAL-PRJ-000034

Photometer Calibrator ICD - HSO-CDF-ICD-013 Issue 3.0

Photometer Calibrator Specifications – HSO-CDF-SP-003 Issue 1.0

Responsible Authority		Signature
Cardiff Product Assurance	Dr I.Walker	
Cardiff SPIRE Management	Dr P.Hargrave	

SECTION 04 - Qualification Status List / Compliance Matrix

Test	Status	Applicable document	Test Institute
Dimension and tolerances to specification	Compliant	PCal ICD – HSO-CDF-ICD-013	UWC
Visual inspection (internal & external)	Passed	Cardiff SPIRE PCal logbook	UWC
Mass	Compliant	PCal ICD - HSO-CDF-ICD-013	UWC
Thermal / vacuum cycles	Passed. Prior to delivery, PCal-PFM underwent a total of five thermal / vacuum cycles to <15-K.	Section 14 & 15 Historical record & Logbook	UWC
Power consumption	Compliant	_	UWC
Vibrations 300K	Passed	AIV-2003-091-VIB, HSO-CDF-RP-078	RAL
Vibrations 4K	Passed	AIV-2003-091-VIB, HSO-CDF-RP-078	RAL
Environmental condition - Vacuum 3x10 ⁻¹ mBar	Compliant	Section 14 & 15 Historical record & Logbook	UWC
Differential pressure (a pumping-out rate of 10mB/sec)	Compliant	Section 14 & 15 Historical record & Logbook	UWC
Pre-bake out (not exceeding 80°C)	Performed & compliant	Section 14 & 15 Historical record & Logbook	UWC
Outgassing	Compliant	By design	UWC
Cleanliness checks, by visual inspection.	Passed	Cardiff SPIRE PCal logbook	UWC
Degradation due to high energy radiation.	Compliant	By design	UWC

Compliance with IRD

Requirement ID	Description	Value	Compliant?
IRD-CALP-R01	Nominal operating output	The area:surface brightness product of the calibrator aperture shall be $\geq 1\%$ of the area:surface brightness product of the telescope image at the position of M4 (with an assumed telescope temperature of 80 K and emissivity of 4%) for $200 < \lambda < 700 \mu m$.	Yes – see test data
IRD-CALP-R02	Operating range	Commandable in 256 steps, with at least 124 steps covering the range from zero output to 2% of the	Yes – 12-bit DAC current drive in range 0-7mA. Required output achieved at 3.9mA.

SPIRE - Photometer Calibration Source (PCal) - PFM - End Item Data Package (EIDP)

		covering the range from zero output to 2% of the power from the telescope	Required output achieved at 3.9mA. 2% of telescope background power at 5.7mA, with 3334 steps from 0 - 5.7mA.
IRD-CALP-R03	Equivalent obscuration of aperture through BSM mirror	The outside envelope of the calibrator housing shall not foul on any part of the BSM for any operational angular position of the BSM	This is the same as IRD-CALP-R09! The old requirement referred to equivalent obscuration of the telescope primary at the BSM pupil
IRD-CALP-R04	Speed of response	Response to step change in applied power, 90% settling time of radiant power output <350ms (70ms)	Compliant. See test data. Rise time constant at nominal power = 96 mS Fall time constant = 52 ms
IRD-CALP-R05	Repeatability	RMS of output signal better than 1% over 20 cycles on to off during a calibration operation of less than 2 minutes. Drift less than10% over lifetime of the mission. Repeatability of signal 1% for 12 calibration operations equi-spaced over a period of 12 hours, with uniform base temperature and drive current.	Yes. See life test report. HSO-CDF-RP-091.
IRD-CALP-R06	Operation	Nominally once per hour for no more than 10 seconds.	Yes
IRD-CALP-R07	Frequency	Continuously or pseudo continuously variable between 0 and 2 Hz.	Yes
IRD-CALP-R08	Interface	The calibrator will be integrated into the beam steering mechanism.	Yes – see PCAL ICD. HSO-CDF-ICD-013-3.0
IRD-CALP-R09	Volume envelope	This shall be compatible with the space available within the BSM enclosure as described in the BSM specification document.	Yes
IRD-CALP-R10	Thermal isolation	The thermal conductance between the calibrator body and the SPIRE optical bench shall be > 2 mW/K.	Determined by BSM structure. Not a requirement on PCal.
IRD-CALP-R11	Operating temperature	<6K	Yes
IRD-CALP-R12	Cold power dissipation	Shall be within the specification given in RD8	RD8 is the now-obsolete SPIRE systems budgets document – SPIRE-ATC-PRJ-000450
IRD-CALP-R13	Warm power dissipation	Requirement removed	
IRD-CALP-R14	Operating voltage	Less than 28V at input power level of 5mW	Yes - see PCAL ICD. HSO-CDF-ICD-013-3.0 Operating voltage at nominal output is 0.9V
IRD-CALP-R15	Redundancy	Cold redundancy for the thermal source	Yes

SPIRE - Photometer Calibration Source (PCal) - PFM - End Item Data Package (EIDP)

Page 13 of 50

IRD-CALP-R16	Lifetime	The calibration source shall be capable of up to 250,000 operational cycles at the nominal electrical power.	Yes. See life test report. HSO-CDF-RP-091.

SECTION 05 - Top Level Drawings (Inc. Family Tree)

Hardware tree

The hardware tree for PCal is shown in Figure 1. Note that the part numbers are given by the general form "PCAL-XXX-nnn", where "XXX" is the model designation (CQM, PFM etc) and "nnn" is the number given in the hardware tree. For instance, the part number for the PCal flight model rear cover would be "PCAL-PFM-202".




Figure 1 Hardware tree for PCal components.

TOP LEVEL DRAWING LIST

Drawing No.	Title

SECTION 06 - Interface Drawings

INTERFACE DRAWING LIST

Drawing No.	Title	Notes
PCAL-ICD-000-B	PCal Interface Control Drawing	This is the flight model interface drawing




Figure 2 Mechanical interface drawing for PCal-PFM

SECTION 07 - Functional, Block & Mechanical Drawings


Component drawings are given in this section.


FUNCTIONAL & BLOCK DRAWING LIST


Drawing No.	Title


MECHANICAL COMPONENT DRAWING LIST


Drawing No.	Title	Notes
SPIRE-PCAL-000-001-2.0	PCAL BODY	
SPIRE-PCAL-000-001-3.0	PCAL SOURCE MOUNT	
SPIRE-PCAL-000-001-4.0	PCAL REAR COVER	
SPIRE-PCAL-000-001-5.0	PCAL SOURCE HOLDER	
SPIRE-PCAL-000-001-6.0	PCAL HARNESS CLAMP	











SECTION 08 – Electrical Circuit Diagrams

The four prime PCAL wires (part of the BSM/PCal prime harness) are terminated on the two prime solder terminals. This harness is duplicated for the redundant systems. The PCAL wiring will consist of an insulated, screened, twisted quad sub-harness. The maximum harness impedance requested for PCAL is 10 Ohms per wire. A schematic of the BSM prime connector wiring is shown in Figure 3. The redundant wiring will be an exact copy of this harness.




Figure 3 Schematic of pin allocations for PCal/BSM prime and redundant connectors. Both prime and redundant connectors are MDM-37 way socket connectors, with jackposts.

Table 1 Pin allocations for PCal-PFM electrical connectors (refer also to ICD – HSO-CDF-ICD-013).

BSM Prime connector Pin#	Label (in accordance with Harness Definition Document)	Function
28	Prime I+_A	Current drive
29	Prime IA	Current return
10	Prime I+_B	Voltage sense +
11	Prime IB	Voltage sense -
BSM Redundant connector Pin#	Label (in accordance with Harness Definition Document)	Function
28	Red I+ A	Current drive
29	Red IA	Current return
10	Red I+_B	Voltage sense +
11	Red IB	Voltage sense -

SECTION 09 - As Built Configuration Items Status List

Files are located in the directories indicated below, and on the accompanying CD-ROM.

Item	Reference	Location	Notes
PCal drawings and manufacturing files		\\Darkstar\Astroworld\Projects\SPIRE\Cardiff_workpackages\Configur ed_Documents\PCAL\PFM-build\Drawings	
Material certificates of conformance		Available at Cardiff for inspection	
Calibration data files		\\Darkstar\Astroworld\Projects\SPIRE\Cardiff_workpackages\Configur ed_Documents\PCAL\PFM-build\Calibration files	
Test data		\\Darkstar\Astroworld\Projects\SPIRE\Cardiff_workpackages\Configur ed_Documents\PCAL\PFM-build\Test data	
Test reports		\\Darkstar\Astroworld\Projects\SPIRE\Cardiff_workpackages\Configur ed_Documents\PCAL\PFM-build\Test reports	
Calibration analysis worksheet		\\Darkstar\Astroworld\Projects\SPIRE\Cardiff_workpackages\Configur ed_Documents\PCAL\PFM-build\Calibration files	

Part number	Description	Details	C. of C. / Batch number
PCAL-PFM-100	HBI-7 – PCal PFM source assembly	Mounting ring manufactured and gold-plated at Cardiff. Source assembled by Haller-Beeman Associates	HBI-7
PCAL-PFM-201	PCAL BODY	Electro-mec (Reading) LTD,	24452 – 17/05/04
PCAL-PFM-202	PCal rear cover	28 Portman Rd,	24452 – 17/05/04
PCAL-PFM-203	PCal harness clip	Reading, Berks RG30 1EA	24452 – 17/05/04
PCAL-PFM-204	Stand offs	Wearnes-Cambion LTD. Part # CA 572-4882-01-05-16	0240025 – 08/11/02
PCAL-PFM-205	2-56 x 3/8 skt cap bolts 2-56 x 1/4 csk bolts	Precision Technology Supplies LTD, The Birches Industrial Estate, Imberhorne Lane, East Grinstead, West Sussex. RH19 1XZ	17010 26469

SPIRE - Photometer Calibration Source (PCal) - PFM - End Item Data Package (EIDP)	Page 26 of 50
of the -1 hotometer datibilition doubte (1 dat) -1 1 W - End term bata 1 dekage (Elbi)	Fage 20 01 30

SECTION 10 – Serialised Components List

As section 9.

SECTION 11 - List of Waivers

No Waivers

Waiver Number	Title	Status

SECTION 12 - Copies of Waivers

N/A

SECTION 13 - Operations Manual

Warm functional test procedure

PCal must NEVER be operated with SPIRE warm (>50K)!

Impedances may be checked at room temperature using either the warm electronics or a multimeter connected in a four-wire configuration to the relevant device lines indicated in section 8. Drive current MUST be limited to 50 micro-Amps. Measured impedance values at room temperature (~300K), 77K and 4.2K are shown for reference.

	Impedance (Ω)				
	Room Temperature	77K	4K		
Prime Source	212.3	201.6	202.1		
Redundant Source	194.9	184.9	185.0		

PCal PFM Specifications as delivered

Specification	Related Requirement	Value	Comments
Thermal source ID		HBI-7	Sapphire substrate, brass wires, indium soldered
Thermal source redundancy	IRD-CALP-R15	Prime and redundant sources are provided	
Prime source impedance – 300K		212.3 Ω	
Prime source impedance – 77K		201.6 Ω	
Prime source impedance – 4K		202.1 Ω	
Redundant source impedance – 300K		194.9 Ω	
Redundant source impedance – 77K		184.9 Ω	
Redundant source impedance – 4K		185.1 Ω	
Prime source 90% rise time	IRD-CALP-R04	96 ms	Value for 3.1 mW applied power. Time constant varies as a function of applied power.
Prime source 90% fall time	IRD-CALP-R04	52 ms	Value for 3.1 mW applied power. Time constant varies as a function of applied power.
Redundant source 90% rise time	IRD-CALP-R04	100 ms	Value for 3.1 mW applied power. Time

			constant varies as a function of applied power.
Redundant source 90% fall time	IRD-CALP-R04	54 ms	Value for 3.1 mW applied power. Time constant varies as a function of applied power.
Power dissipation at nominal operating output	IRD-CALP-R01, IRD- CALP-R12	3.1 mW	
Current drive for nominal operating output		Prime - 3.9 mA Redundant – 4.03 mA	
Nominal current range	IRD-CALP-R02	Prime – 0-4.9 mA Redundant – 0-5.2 mA	This current range is for a maximum power dissipation of 5mW
Absolute maximum current rating	IRD-CALP-R02	7 mA	This current level will dissipate 9.9 mW in the PCal emitter (prime)
Nominal operating voltage	IRD-CALP-R14	Prime - 0.79 V Redundant – 0.75 V	Voltage at nominal output
Voltage range		Prime – 0-1.41 V Redundant – 1.35 V	This voltage range is for a maximum power dissipation of 9.9 mW (at 7 mA drive)
Stability & repeatability	IRD-CALP-R05	RMS better than 1% over lifetime. Cycle-to-cycle repeatability better than 1%.	Two identical sources (four emitters) have undergone 375,000 cycles each (5mW applied) without any degradation or variation in the performance of the devices.
Frequency	IRD-CALP-R07	Prime – 96 ms Redundant – 100 ms	Time constant at nominal power. Time constant varies as a function of applied power from 66.6 ms to 125 ms (1mW to 5mW)
Operating temperature	IRD-CALP-R11	<10K	PCal must not be operated above 10K ambient, apart from continuity checks (10 micro-Amp maximum drive)
Lifetime	IRD-CALP-R16	>375,000 cycles	At 5mW power dissipation per cycle
Mass		28.9g	<u> </u>
Volume	IRD-CALP-R09		The volume envelope for PCal is shown in the ICD – HSO-CDF-ICD-013 issue 3.0

SECTION 14 - Historical Record

Summary. See section 15 – Logbook / Diary of events – for detailed record.

Note – "Thermal / Vacuum cycle" means one cycle 300K – 4K – 300K under vacuum in the test dewar.

Date	Event	Notes		
15/07/04	HBI-7 source received from Haller-Beeman associates			
27/07/04	Visual inspection and impedance checks	Carried out by P. Hargrave. Inspection_test_report_HBI-1_HBI-8.doc		
15/09/04	HBI-7 source mounted in DM structure for photometric testing	g P. Hargrave. Electrical wires left over-length for this installation. Wires will		
		be cut to length and soldered to new stand-offs prior to delivery.		
17/09/04	Thermal / vacuum cycle, photometric calibration	I. Didschuns. HBI-7_Results-34.doc.		
15/11/04	Selection of HBI-7 source as flight model source			
17/11/04	HBI-7 source mounted in flight model structure	P. Hargrave. Extended wires left in place for photometric tests.		
23/11/04	Thermal / vacuum cycle, final photometric calibration as PFM assembly	I. Didschuns. HBI-7_Results-39.doc		
06/12/04	PCal-PFM assembly removed from test dewar & placed in	P. Hargrave.		
	storage.			
09/02/05	New stand-offs installed and wires cut to final length.	P. Hargrave		
02/02/05	Thermal / vacuum cycles (2 off)	P. Hargrave. Impedances monitored before & after.		
14/02/05	Visual inspection & final packing.	P. Hargrave		

SECTION 15 - Logbook / Diary of Events

See historical record.

Laboratory logbook "PCAL LOG #1" stored securely at Cardiff University.

See also "X:\Cardiff_workpackages\Configured_Documents\PCAL\PFM-build\Test reports\Inspection_test_report_HBI-1_HBI-8.doc".

SECTION 16 - Operating Time / Cycle Record

See also section 14 and section 15.

PCal-PFM has undergone a total of four thermal / vacuum cycles 300K – 4K – 300K.

While at 4K (baseplate temperature), the PCal sources have been heated to varying temperatures, by the application of currents in the range 0 – 7.0 mA, a number of times as indicated in section 14 and the associated test reports. It is estimated that the total number of "flashes" carried out on each source is less than 200.

SECTION 17 – Connector Mating Record – N/A

Connector M	ate / Demate log	3							
Project		SPIRE		Experiment					
Subsystem				Model	PFM	ID			
ID	ID		ID		ID		ID		
Mate date	Demate date	Mate date	Demate date	Mate date	Demate date	Mate date	Demate date	Mate date	Demate date
After 5 cycles Connect I/D	After 5 cycles, carry out visual inspection. Record result below. Connect I/D Debris Bent pins Remarks Pass Fail Signature								
ID ID			ID		ID		ID		
Mate date	Demate date	Mate date	Demate date	Mate date	Demate date	Mate date	Demate date	Mate date	Demate date
	es, carry out vis	<u> </u>	n with magnifica		esult below.	Desc	L Fail	Cianatura	
Connect I/D		Debris	Bent pins	Remarks		Pass	Fail	Signature	
NOTE: In cas	se of failure, an	NCR is require	ed.						

SECTION 18 - Age Sensitive Items Record

There are no age-sensitive items.

SECTION 19 – Pressure Vessel History / Test Record

N/A

SECTION 20 - Calibration Data Record

The calibration data in this section are presented in terms of the requirement on photometric output, and plotted as Figure 4. The photometric output in this plot is expressed in terms of the area:surface-brightness product of the PCal aperture, as a fraction of the area:surface-brightness product of the telescope image at the position of M4. The required output (>1% of telescope image) is achieved with a power dissipation of 1.7mW (2.59mA).




Figure 4 Area: surface-brightness product (ASBP) of PCal aperture as a fraction of the ASBP of the telescope image at M4, plotted as a function of applied power.




Figure 5 Area: surface-brightness product (ASBP) of PCal aperture as a fraction of the ASBP of the telescope image at M4, plotted as a function of drive current (prime source).




Figure 6 Area: surface-brightness product (ASBP) of PCal aperture as a fraction of the ASBP of the telescope image at M4, plotted as a function of drive current (redundant source).

SECTION 21 - Temporary Installation Record

See section 14 – Historical record.

The HBI-7 source was temporarily installed in the PCAL-DM structure for the initial photometric tests (test #34) on 15/09/04.

The full PCal-PFM assembly (flight source in flight structure) was installed in the test cryostat on 17/11/04, with extended wires.

The final PCal-PFM assembly (new stand-offs, wires cut to length) was installed in the test dewar on 02/02/05 for final thermal / vacuum cycling, after which impedance & visual checks were made.

SECTION 22 - Open Work / Deferred Work / Open Tests

Absolute calibration check. The flight-spare PCal assembly will be calibrated directly against a cryogenic black-body source, using a 300mK bolometer. This work is surplus to the delivery requirements. There is no requirement for absolute calibration.

SECTION 23 - List of Non-Conformance Reports

None.

SECTION 24 - Copies of Non-Conformance Reports

N/A

SECTION 25 - Test Reports

See appendices.

SECTION 26 – Assembly Record

N/A

SECTION 27 - Reference List of EIDP's

None

Associated

<u>Title</u> (Listed in alphabetical order)	ID (Serial No.)	<u>Acronym</u>	Document No.	<u>Issue</u>	<u>Date</u>

Lower Level

Title (Listed in alphabetical order)	ID (Serial No.)	Acronym	Document No.	<u>Issue</u>	<u>Date</u>

SECTION 28 - Mass Records

Assembly	Final measured mass		
PCal-PFM assembly	28.9g		

SECTION 29 - Cleanliness Statement – will be signed once final approved version PDF'd Statement

The PFM Photometer Calibration Source assembly (PCAL-PFM-000) has been cleaned, assembled and tested within a class 1000 clean room to meet the requirements of the Cardiff PA plan (HSO-CDF-PL-007).

Signed	Peter Hargrave, Technical Manager, Cardiff-SPIRE deliverables.
Signed	lan Walker, Programme Manager, Cardiff AIG.
Date	

Extra Information

A dedicated Herschel-Planck clean room is available in the Cardiff AIG labs, class 1 000, with class 100 laminar flow cabinets.

For cooldown tests (thermal cycles) the PFM assembly was integrated to the Cardiff test dewar within the clean room annex (approx. Class 10,000 – exposure ~15 minutes per thermal cycle).

The PFM assembly was also integrated to the RAL cold vibration test facility. For the duration of this test, the Scal-PFM assembly was kept in a sealed bag with vent holes.

ECTION 30 - Other Useful Information notos will go here once I've downloaded them			
SPIRE - Photometer Calibration Source (PCal) - PFM - End Item Data Package (EIDP)	Page 40 of 50		

SECTION 31 - DPL/DML

Refer to the Cardiff-SPIRE PFM deliverables lists.

Cardiff-SPIRE-DML	HSO-CDF-LI-074
Cardiff-SPIRE-DMPL	HSO-CDF-LI-075
Cardiff-SPIRE-DPL	HSO-CDF-LI-076

SECTION 32 – List of Appendices/Attachments

Appendix #	Title (Listed in alphabetical order)	Document No.	<u>Issue</u>	<u>Date</u>	<u>Notes</u>
Α	PCal-PFM assembly test report	HSO-CDF-RP-094	1.0		HBI-7_Results-39-HSO-CDF-RP-094.doc HBI-7-39.doc
В	HBI-7 test report	HSO-CDF-RP-093	1.0		HBI-7_Results-34-HSO-CDF-RP-093.doc HBI-7-34.doc
С	Reports from Cryogenic Vibration Tests	HSO-CDF-RP-078	1.0		cold-shake2-rpt-HSO-CDF-RP-078.doc
		AIV-2003-091-VIB	1.0		AIV-2003-091-VIB SPIRE CARDIFF COMPONENTS.pdf
D	HBI-8 lifetest report	HSO-CDF-RP-091	1.0		HBI-8 life test-Results-HSO-CDF-RP-091.doc
	HBI-8 lifetest procedure	HSO-CDF-PR-092	1.0		Test procedure and record-HBI-8-HSO-CDF- PR-092.doc Flight replica device.
E	HB-27 lifetest report	HSO-CDF-RP-067	1.0		HB27 life test-Results-HSO-CDF-RP-067.doc Small difference from flight device.
F	PCal FMECA report	HSO-CDF-RP-032	2.0		Pcal_FMECA_HSO_CDF_RP_032_2_0.doc
G	PCal Interface control document	HSO-CDF-ICD-013	3.0		PCAL_ICD_HSO_CDF_ICD_013_3_0.doc
Н	Inspection & test report – PCal candidate sources	HSO-CDF-RP-095	1.0		Inspection_test_report_HBI-1_HBI-8.doc
I	Copies of certificates of conformance				

Appendix A PCal-PFM Assembly test report

Appendix B
HBI-7 test report

Appendix C Reports from cryogenic vibration tests

Appendix D PCal-PFM replica lifetest report and procedure

Appendix E HB-27 lifetest report

Appendix F PCal FMECA report

Appendix G
PCal-PFM ICD

Appendix H Inspection & test report – PCal candidate sources

Appendix H Inspection & test report – PCal candidate sources