

PACS\_CS\_SFT\_Cold\_OBS  
 File: H\_COP\_PAC\_L001.xls  
 Author: R. Biggins



## Procedure Summary

### Objectives

The objective of this procedure is to execute a short functional test on the Calibration Source (CS)

Based on procedure:  
 PACS\_CS\_SFT\_Cold\_OBS (v1)

### Summary of Constraints

This procedure should be executed as part of the Short Functional Test (HeII conditions)

RT Science must be enabled to receive the Diagnostic HK packets

### Spacecraft Configuration

#### Start of Procedure

PACS in NO\_PRIME (SAFE) mode  
 - PACS is generating Spectroscopy HK

#### End of Procedure

PACS in NO\_PRIME (SAFE) mode  
 - PACS is generating Spectroscopy HK

### Reference File(s)

#### Input Command Sequences

#### Output Command Sequences

HCPL001

### Referenced Displays

ANDs	GRDs	SLDs
ZAZ98999	PG007420	
PA019420		
PA032420		

### Configuration Control Information


DATE	FOP ISSUE	VERSION	MODIFICATION DESCRIPTION	AUTHOR	SPR REF
12/08/08		1	Created	R. Biggins	
13/11/08	2	2	Updates due to initial testing - New step (1) added for prerequisite check - Initial OBSID value changed to FP - Time tags added to commands in step 3	R. Biggins	
15/04/09	2.3	3	Final updates before flight - Summary updated - TC flags updated	R. Biggins	

Status : Version 3 - Unchanged  
 Last Checkin: 15/04/09

PACS\_CS\_SFT\_Cold\_OBS  
File: H\_COP\_PAC\_L001.xls  
Author: R. Biggins



## Procedure Flowchart Overview





PACS\_CS\_SFT\_Cold\_OBS  
 File: H\_COP\_PAC\_L001.xls  
 Author: R. Biggins



Step No.	Time	Activity/Remarks	TC/TLM	Display/ Branch
		DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_CHECKSUM PP066420  TC Control Flags :  GBM IL DSE --Y -- --  Subsch. ID : 90 Det. descr. : WRITE THE LIST OF DIAGNOSTIC HK INTO DMC MEMORY	211 <hex> 21E <hex> 21F <hex> 220 <hex> 221 <hex> FFFF <hex> F109 <hex>	
	ET=+00.00.01 UT=+00.00.01	DMC_START_DIAG_HK  Command Parameter(s) : DIAG_HK_PERIOD PP076420  Subsch. ID : 90 Det. descr. : START THE ACQUISITION OF THE DIAGNOSTIC HK	DMC_START_DIAG_HK PC146420  300 <dec>	
		<b>NOTE:</b> The verification of the generation of the <b>TM(21,3)</b> diagnostic packets cannot be done on the MCS, but can be seen on the NCTRS (VC7)		
	ET=+00.00.01 UT=+00.00.01	Execute Telecommand  DMC_WRT_CS1_CONF_PAR  Command Parameter(s) : DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_CHECKSUM PP066420  TC Control Flags :  GBM IL DSE  --Y -- --	PC166420  F4240 <hex> 1388 <hex> CCD <hex> 741 <hex> 147 <hex> 7FFF <hex> 0 <hex> 5C9F <hex>	
		Subsch. ID : 90 Det. descr. : WRITE THE PARAMETERS TO CONFIGURE THE CS1 TEMP REGULATOR		

PACS\_CS\_SFT\_Cold\_OBS  
 File: H\_COP\_PAC\_L001.xls  
 Author: R. Biggins



Step No.	Time	Activity/Remarks	TC/TLM	Display/ Branch
	ET=+00.00.01 UT=+00.00.01	Execute Telecommand  DMC_WRT_CS2_CONF_PAR  Command Parameter(s) : DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_4_BYTES_WORDS_DATA PP067420 DMC_CHECKSUM PP066420  Subsch. ID : 90 Det. descr. : WRITE THE PARAMETERS TO CONFIGURE THE CS2 TEMP REGULATOR	PC167420   F4240 <hex> 1388 <hex> CCD <hex> 741 <hex> 147 <hex> 7FFF <hex> 0 <hex> 5C9F <hex>	
	ET=+00.00.01 UT=+00.00.01	DMC_SWON_BB_1_CONT  Subsch. ID : 90 Det. descr. : SWITCH ON BLACK BODY 1 CONTROLLER	PC138420	
	ET=+00.00.01 UT=+00.00.01	DMC_ENABLE_BB_1_CONT  Subsch. ID : 90 Det. descr. : ENABLE BB1 CONTROLLER	PC206420	
	ET=+00.00.00 UT=+00.00.00	DMC_SET_TEMP_BB_1  DMC_SET_TEMP_BB_1  Command Parameter(s) : DMC_BB_TEMP PP094420  Subsch. ID : 90 Det. descr. : SET THE TEMPERATURE OF THE BLACK BODY 1	PC140420  80.0 Ohm	
	ET=+00.00.01 UT=+00.00.01	DMC_SWON_BB_2_CONT  Subsch. ID : 90 Det. descr. : SWITCH ON BLACK BODY 2 CONTROLLER	PC142420	
	ET=+00.00.01 UT=+00.00.01	DMC_ENABLE_BB_2_CONT  Subsch. ID : 90 Det. descr. : ENABLE BB2 CONTROLLER	PC208420	
	ET=+00.00.01 UT=+00.00.01	DMC_SET_TEMP_BB_2  DMC_SET_TEMP_BB_2  Command Parameter(s) : DMC_BB_TEMP PP094420  Subsch. ID : 90 Det. descr. : SET THE TEMPERATURE OF THE BLACK BODY 2	PC144420  92.0 Ohm	

PACS\_CS\_SFT\_Cold\_OBS  
 File: H\_COP\_PAC\_L001.xls  
 Author: R. Biggins



Step No.	Time	Activity/Remarks	TC/TLM	Display/ Branch
		Verify Telemetry DM_CS1_TARGET PM263420	= 80.0 Ohm	AND=PA019420
		Verify Telemetry DM_CS2_TARGET PM264420	= 92.0 Ohm	AND=PA019420
		Verify Telemetry DM_CS1C_POWER PM544420	= POWER ON	AND=PA032420
		Verify Telemetry DM_CS2C_POWER PM558420	= POWER ON	AND=PA032420
		During the time until the last TC is sent (10 minutes), the following resistor values should increase from Liquid Helium values towards the defined set points. Note that the CS will not settle on the target values as it would take too long for the SFT.		
		Verify Telemetry DM_CS1_RES_VAL PM249420	= 80.0 Ohm	GRD=PG007420
		Verify Telemetry DM_CS2_RES_VAL PM251420	= 92.0 Ohm	GRD=PG007420
	ET=+00.10.00 UT=+00.10.00	DMC_STOP_DIAG_HK DMC_STOP_DIAG_HK TC Control Flags : Subsch. ID : 90 Det. descr. : STOP THE ACQUISITION OF THE DIAGNOSTIC HK	GBM IL DSE --Y -- ---	PC147420
		<b>NOTE:</b> The verification of the end of generation of the <b>TM(21,3)</b> diagnostic packets cannot be done on the MCS, but can be seen on the NCTRS (VC7)		
4		Set default variables		Next Step: END
	ET=+00.00.01 UT=+00.00.01	DMC_SET_OBSID DMC_SET_OBSID Command Parameter(s) : OBSERVATION_ID PP069420 TC Control Flags : Subsch. ID : 90 Det. descr. : SET OBSID IN DMC	GBM IL DSE --Y -- ---	PC078420 00000000 <hex>

PACS\_CS\_SFT\_Cold\_OBS  
 File: H\_COP\_PAC\_L001.xls  
 Author: R. Biggins



Step No.	Time	Activity/Remarks	TC/TLM	Display/ Branch
	ET=+00.00.00 UT=+00.00.00	DMC_SET_BBID  DMC_SET_BBID  Command Parameter(s) : BUILDING_BLOCK_ID PP070420  Subsch. ID : 90 Det. descr. : SET BBID IN DMC	PC079420  40000000 <hex>	
		Verify Telemetry  DM_OBSID PM028420	= 00000000 <hex>	AND=ZAZ98999
		Verify Telemetry  DM_BBID PM029420	= 40000000 <hex>	AND=ZAZ98999
<b>End of Procedure</b>				