

# AMIE FM Calibration - list of activities

Reference: S1-AMIE-SSD-LI-001

Issue: 1 Rev.: -

Date: 03 Dec 2001

Page: 1

Date/Time	Filename(s)	Measurement name (flat fielding, dark...)	Temperature of CCD	"Ambient" temp i.e. chamber	Illumination (int. sphere, diffusor, none)	Comments
2001.12.04 19h20	2001.12.04-AMIE-FM-Dark -30°C-10...ms.pdf	Dark	-30°C	-30°C	None	Images for each IT
	10, 50, 100, 200, 500, 1000, 2000, 4000, 8000, 16k, 32k, 64k	Dark	-20°C	-29°C	None	
2001.12.05		Dark	-10°C	-19°C	None	
		Dark	0°C	-9°C	None	
			10°C	1°C		
14h			20°C	10,5°C		
15h10			30°C	19,2°C		
			40°C	29,4°C		
			50°C	39,8°C		
		Flat fields	50°C	41,6°C	white source + diffusor	
			40°C	30°C		
			30°C	24°C		

